


7470A

LB104015

INSTRUMENT ID: CU1

Linear



A= 0.0000e+000

B= 2.5291e-004

C= -1.5149e-002

Rho= 0.9994941

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	26	0.000	26				
0.2	0.200	0.196	-0.004	834	0.0 %	834				
2.5	2.500	2.625	0.125	10441	0.0 %	10441				
5.0	5.000	4.961	-0.039	19674	0.0 %	19674				
7.5	7.500	7.287	-0.213	28871	0.0 %	28871				
10.0	10.000	10.140	0.140	40155	0.0 %	40155				

LB104015 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 17 Jul 2019 10:07:50					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	26	-	0.0000	7470A	PPB	17 Jul 2019 10:11:48	S	Std
0.2 - 1	50-2	834	-	0.2000	7470A	PPB	17 Jul 2019 10:13:57	S	Std
2.5 - 1	52-5	10441	-	2.5000	7470A	PPB	17 Jul 2019 10:16:06	S	Std
5.0 - 1	55-0	19674	-	5.0000	7470A	PPB	17 Jul 2019 10:18:14	S	Std
7.5 - 1	57-5	28871	-	7.5000	7470A	PPB	17 Jul 2019 10:20:23	S	Std
10.0 - 1	50-0	40155	-	10.0000	7470A	PPB	17 Jul 2019 10:27:30	S	Std
ICV35 - 1	ICV35	15178	3.8235	-	7470A	PPB	17 Jul 2019 10:30:05	U	SMPL
ICB35 - 1	ICB35	-10	-0.0177	-	7470A	PPB	17 Jul 2019 10:32:11	U	SMPL
CCV36 - 1	CCV36	19887	5.0144	-	7470A	PPB	17 Jul 2019 10:34:18	U	SMPL
CCB36 - 1	CCB36	103	0.0109	-	7470A	PPB	17 Jul 2019 10:36:25	U	SMPL
CRA - 1	CRA	833	0.1955	-	7470A	PPB	17 Jul 2019 10:38:31	U	SMPL
HighStd - 1	HighStd	41081	10.3745	-	7470A	PPB	17 Jul 2019 10:40:42	U	SMPL
ChkStd - 1	ChkStd	28681	7.2385	-	7470A	PPB	17 Jul 2019 10:42:49	U	SMPL
PB121449BL - 1	PBW	-51	-0.0280	-	7470A	PPB	17 Jul 2019 10:44:57	U	SMPL
PB121449BS - 1	LCSW	16249	4.0943	-	7470A	PPB	17 Jul 2019 10:47:04	U	SMPL
K3620-02 - 1	CL-01-071219-A	-13	-0.0184	-	7470A	PPB	17 Jul 2019 10:49:12	U	SMPL
K3620-02DUP - 1	CL-01-071219-ADUP	23	-0.0093	-	7470A	PPB	17 Jul 2019 10:51:20	U	SMPL
K3620-02MS - 1	CL-01-071219-AMS	16151	4.0695	-	7470A	PPB	17 Jul 2019 10:53:28	U	SMPL
K3620-02MSD - 1	CL-01-071219-AMSD	16146	4.0683	-	7470A	PPB	17 Jul 2019 10:55:36	U	SMPL
K3620-04 - 1	CL-01-071219-B	-24	-0.0212	-	7470A	PPB	17 Jul 2019 10:57:44	U	SMPL
CCV37 - 1	CCV37	19770	4.9848	-	7470A	PPB	17 Jul 2019 10:59:52	U	SMPL
CCB37 - 1	CCB37	-17	-0.0194	-	7470A	PPB	17 Jul 2019 11:01:58	U	SMPL
K3620-06 - 1	CL-01-071219-C	9	-0.0129	-	7470A	PPB	17 Jul 2019 11:04:05	U	SMPL
K3620-08 - 1	CL-01-071219-D	10	-0.0126	-	7470A	PPB	17 Jul 2019 11:06:12	U	SMPL
K3620-10 - 1	CL-01-071219-E	20	-0.0101	-	7470A	PPB	17 Jul 2019 11:08:18	U	SMPL
PB121450BL - 1	PBW	16	-0.0111	-	7470A	PPB	17 Jul 2019 11:10:25	U	SMPL
PB121450BS - 1	LCSW	16306	4.1087	-	7470A	PPB	17 Jul 2019 11:12:32	U	SMPL
K3622-06 - 1	OR-03-071519-A	-7	-0.0169	-	7470A	PPB	17 Jul 2019 11:14:40	U	SMPL
K3622-06DUP - 1	OR-03-071519-ADUP	14	-0.0116	-	7470A	PPB	17 Jul 2019 11:16:47	U	SMPL
K3622-06MS - 1	OR-03-071519-AMS	16354	4.1209	-	7470A	PPB	17 Jul 2019 11:18:55	U	SMPL
K3622-06MSD - 1	OR-03-071519-AMSD	15937	4.0154	-	7470A	PPB	17 Jul 2019 11:21:04	U	SMPL
K3622-08 - 1	OR-03-071519-B	-5	-0.0164	-	7470A	PPB	17 Jul 2019 11:23:10	U	SMPL
CCV38 - 1	CCV38	19493	4.9148	-	7470A	PPB	17 Jul 2019 11:25:17	U	SMPL
CCB38 - 1	CCB38	-31	-0.0230	-	7470A	PPB	17 Jul 2019 11:27:23	U	SMPL
K3623-02 - 1	EO-02-071519-A	7	-0.0134	-	7470A	PPB	17 Jul 2019 11:29:31	U	SMPL
K3623-04 - 1	EO-02-071519-B	18	-0.0106	-	7470A	PPB	17 Jul 2019 11:31:37	U	SMPL
K3623-06 - 1	EO-02-071519-C	20	-0.0101	-	7470A	PPB	17 Jul 2019 11:33:44	U	SMPL
K3627-04 - 1	TR-04-071519	14	-0.0116	-	7470A	PPB	17 Jul 2019 11:35:51	U	SMPL
K3820-02 - 1	RBR-200050	4	-0.0141	-	7470A	PPB	17 Jul 2019 11:37:58	U	SMPL
K3826-02 - 1	RBR-200034	21	-0.0098	-	7470A	PPB	17 Jul 2019 11:40:06	U	SMPL
K3826-04 - 1	RBR-200036	16	-0.0111	-	7470A	PPB	17 Jul 2019 11:42:15	U	SMPL
K3829-01 - 1	IDW-EW03-TRENCH	142	0.0208	-	7470A	PPB	17 Jul 2019 11:44:32	U	SMPL
K3829-02 - 1	IDW-EW03-CONCRETE	-6	-0.0167	-	7470A	PPB	17 Jul 2019 11:46:41	U	SMPL
K3830-02 - 1	RT-3456	8	-0.0131	-	7470A	PPB	17 Jul 2019 11:48:48	U	SMPL
CCV39 - 1	CCV39	19762	4.9828	-	7470A	PPB	17 Jul 2019 11:53:36	U	SMPL
CCB39 - 1	CCB39	-8	-0.0172	-	7470A	PPB	17 Jul 2019 11:55:43	U	SMPL
K3830-04 - 1	RT-4651	6	-0.0136	-	7470A	PPB	17 Jul 2019 11:57:50	U	SMPL
K3830-06 - 1	RT-1873	24	-0.0091	-	7470A	PPB	17 Jul 2019 11:59:56	U	SMPL
K3830-08 - 1	RBR-200027	17	-0.0108	-	7470A	PPB	17 Jul 2019 12:02:04	U	SMPL
PB121449TB - 1	PB121353TB	31	-0.0073	-	7470A	PPB	17 Jul 2019 12:04:11	U	SMPL
PB121450TB - 1	PB121397TB	28	-0.0081	-	7470A	PPB	17 Jul 2019 12:06:18	U	SMPL
K3620-02LX5 - 1		4	-0.0141	-	7470A	PPB	17 Jul 2019 12:08:26	U	SMPL
K3620-02A - 1		16285	4.1034	-	7470A	PPB	17 Jul 2019 12:10:34	U	SMPL
K3622-06LX5 - 1		-9	-0.0174	-	7470A	PPB	17 Jul 2019 12:12:42	U	SMPL
K3622-06A - 1		16257	4.0964	-	7470A	PPB	17 Jul 2019 12:14:51	U	SMPL
CCV40 - 1	CCV40	20314	5.1224	-	7470A	PPB	17 Jul 2019 12:16:59	U	SMPL
CCB40 - 1	CCB40	-33	-0.0235	-	7470A	PPB	17 Jul 2019 12:19:07	U	SMPL