



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104018

Slope : 98.3

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	07/17/2019	11:25
2	CAL2	1	Water	NA	NA	21.1	7.00	07/17/2019	11:26
3	CAL3	1	Water	NA	NA	21.1	10.03	07/17/2019	11:27
4	ICV	1	Water	NA	NA	21.1	7.01	07/17/2019	11:30
5	CCV1	1	Water	NA	NA	21.2	2.02	07/17/2019	11:31
6	K3854-01	1	Water	NA	NA	18.6	6.72	07/17/2019	11:32
7	K3854-01DUP	1	Water	NA	NA	18.7	6.73	07/17/2019	11:34
8	CCV2	1	Water	NA	NA	21.2	12.02	07/17/2019	11:35

2-B 104018

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3854 WorkList ID : 128659 Department : Wet-Chemistry Date : 07-17-2019 11:19:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/16/2019	Water	K3854-01	✓ pH	Cool 4 deg C	APT101	M63	WGGM-19-WWW-TSF-028	07/16/2019	SM 4500-H B

Date/Time 07-17-19 11:20
Received by: fr
Relinquished by: es

Date/Time 07-17-19 12:00
Received by: CS
Relinquished by: fr