



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104028

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/17/2019	16:00
2	CAL2	1	Water	NA	NA	20.2	7.01	07/17/2019	16:01
3	CAL3	1	Water	NA	NA	20.2	10.03	07/17/2019	16:02
4	ICV	1	Water	NA	NA	20.2	6.99	07/17/2019	16:04
5	CCV1	1	Water	NA	NA	20.1	2.02	07/17/2019	16:05
6	K3857-02	1	Water	NA	NA	27.2	7.51	07/17/2019	16:06
7	K3857-02DUP	1	Water	NA	NA	27.3	7.52	07/17/2019	16:07
8	CCV2	1	Water	NA	NA	20.1	12.01	07/17/2019	16:10

28104028

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K3867 WorkList ID : 128669 Department : Wet-Chemistry Date : 07-17-2019 15:58:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/17/2019	Water	K3857-02	Corrosivity	Cool 4 deg C	PSEG01	M52	36916	07/17/2019	9040C

Date/Time 07-17-14 16:00
Received by: Jn
Relinquished by: cn

Date/Time 07-17-19 17:00
Received by:
Relinquished by: Jn