



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB104035

ANALYST: chirag
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC PH METER-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include various calibration standards and reagents like Hexavalent Chromium ICV Std, Hexavalent Chromium CCV Std, and 5N sulfuric acid.

Intercept: 0.0003 Slope: 0.8001 Regression: 0.999997

Main data table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: chirag

Parameter: Hexavalent Chromiu

SUPERVISOR REVIEW BY: apatel

Run Number: LB104035

pH Meter ID: WC PH METER-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH		Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
						HN03	H2SO4	Backgrnd	Color				
1	ICV	0.5	1	100	100		1.65	0.000	0.410	0.410	0.512	07/17/2019	12:40
2	ICB		1	100	100		2.10	0.000	0.000	0.000	0.000	07/17/2019	12:41
3	CCV1	0.5	1	100	100		1.65	0.000	0.406	0.406	0.507	07/17/2019	13:03
4	CCB1		1	100	100		1.60	0.000	0.000	0.000	0.000	07/17/2019	13:04
5	PB121458BL		1	2.53	100	7.25	1.56	0.000	0.000	0.000	0.000	07/17/2019	13:05
6	PB121458BS	20	1	2.52	100	7.65	2.01	0.000	0.362	0.362	0.452	07/17/2019	13:06
7	K3840-01		1	2.53	100	7.12	1.65	0.013	0.013	0.000	0.000	07/17/2019	13:07
8	K3840-01DUP		1	2.51	100	7.85	2.05	0.013	0.013	0.000	0.000	07/17/2019	13:08
9	K3840-01MSPre	40	2	2.50	100	7.15	1.56	0.013	0.202	0.189	0.236	07/17/2019	13:08
10	K3840-01MS2Ins	1284	40	2.53	100	7.65	1.65	0.013	0.648	0.635	0.793	07/17/2019	13:09
11	K3840-01MS3Post	40	2	2.52	100	7.12	1.69	0.000	0.315	0.315	0.393	07/17/2019	13:10
12	K3840-03		1	2.55	100	7.50	1.89	0.025	0.025	0.000	0.000	07/17/2019	13:11
13	K3852-01		1	2.52	100	7.25	1.56	0.013	0.013	0.000	0.000	07/17/2019	13:12
14	K3852-02		1	2.54	100	7.85	2.05	0.018	0.018	0.000	0.000	07/17/2019	13:13
15	CCV2	0.5	1	100	100		1.65	0.000	0.399	0.399	0.498	07/17/2019	13:14
16	CCB2		1	100	100		2.01	0.000	0.000	0.000	0.000	07/17/2019	13:15
17	K3852-03		1	2.54	100	7.15	1.56	0.022	0.022	0.000	0.000	07/17/2019	13:16
18	CCV3	0.5	1	100	100		2.10	0.000	0.401	0.401	0.501	07/17/2019	13:17
19	CCB3		1	100	100		1.56	0.000	0.000	0.000	0.000	07/17/2019	13:18