



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: chirag

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB104041

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP76219
COD calibration std. 100 ppm	WP76220
COD calibration std. 10 ppm	WP76221
COD calibration std. 150 ppm	WP76222
COD calibration std. 50 ppm	WP76223
COD calibration std. 0 ppm	WP76224
COD CCV std, 50ppm	WP76225
COD Digestion Vials Low Level 0-150Mg/L	W2528
COD ICV-LCS std, 50ppm	WP76446
COD LOD std, 5ppm	WP76447

Temp In (C): 150	Date In: 07/18/2019	Time In: 08:39
Temp Out (C): 150	Date Out: 07/18/2019	Time Out: 10:39

Intercept: 0.2594

Slope: 1.0023

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	07/11/2019	12:00
2	CAL2	10	1	Water	11.000	07/11/2019	12:01
3	CAL3	50	1	Water	50.000	07/11/2019	12:02
4	CAL4	100	1	Water	100.000	07/11/2019	12:03
5	CAL5	150	1	Water	151.000	07/11/2019	12:04

Analysis Method: SM5220 D

ANALYST: chirag

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB104041

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.626	07/11/2019	12:05
2	ICB				1	Water	0.000	-0.259	07/11/2019	12:06
3	CCV1	50			1	Water	51.000	50.624	07/18/2019	11:16
4	CCB1				1	Water	0.000	-0.259	07/18/2019	11:16
5	LB104041BL				1	Water	0.000	-0.259	07/18/2019	11:17
6	LB104041BS	50			1	Water	49.000	48.629	07/18/2019	11:18
7	K3591-09				1	Water	4.000	3.732	07/18/2019	11:18
8	K3838-03				10	Water	34.000	33.663	07/18/2019	11:19
9	K3856-02				1	Water	3.000	2.734	07/18/2019	11:19
10	K3856-03				1	Water	3.000	2.734	07/18/2019	11:20
11	K3857-01				1	Water	30.000	29.672	07/18/2019	11:20
12	K3859-01				1	Water	4.000	3.732	07/18/2019	11:21
13	K3859-02				1	Water	0.000	-0.259	07/18/2019	11:21
14	K3859-02DUP				1	Water	0.000	-0.259	07/18/2019	11:22
15	CCV2	50			1	Water	50.000	49.626	07/18/2019	11:23
16	CCB2				1	Water	0.000	-0.259	07/18/2019	11:23
17	K3859-02MS	50			1	Water	50.000	49.626	07/18/2019	11:24
18	K3859-02MSD	50			1	Water	50.000	49.626	07/18/2019	11:24
19	CCV3	50			1	Water	49.000	48.629	07/18/2019	11:25
20	CCB3				1	Water	0.000	-0.259	07/18/2019	11:25