


LB104049

ISM02.4

INSTRUMENT ID: C01



Linear

A= 0.0000e+000

B= 2.4658e-004

C= -1.4955e-002

Rho= 0.9999314

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.008	0.008	95	0.000	95				
0.05	0.050									
0.20	0.200	0.180	-0.020	789	0.0 %	789				
2.50	2.500	2.465	-0.035	10056	0.0 %	10056				
5.00	5.000	5.092	0.092	20710	0.0 %	20710				
7.50	7.500	7.473	-0.027	30366	0.0 %	30366				
10.0	10.000	9.983	-0.017	40544	0.0 %	40544				

20
—
-10
-1
2
0
0

02 07-18-19

LB104049

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	95	-		0 ISM02.4	PPB	7/18/2019 12:37	Std
	0.2 S0.2	789	-		0.2 ISM02.4	PPB	7/18/2019 12:40	Std
	2.5 S2.5	10056	-		2.5 ISM02.4	PPB	7/18/2019 12:42	Std
	5 S5.0	20710	-		5 ISM02.4	PPB	7/18/2019 12:44	Std
	7.5 S7.5	30366	-		7.5 ISM02.4	PPB	7/18/2019 12:46	Std
	10 S10.0	40544	-		10 ISM02.4	PPB	7/18/2019 12:49	Std
ICV37	ICV37	16535	4.0623 -		ISM02.4	PPB	7/18/2019 12:51	SMPL
ICB37	ICB37	33	-0.0068 -		ISM02.4	PPB	7/18/2019 12:54	SMPL
CCV44	CCV44	20371	5.0082 -		ISM02.4	PPB	7/18/2019 12:56	SMPL
CCB44	CCB44	15	-0.0113 -		ISM02.4	PPB	7/18/2019 12:58	SMPL
PB121494BL	PBW	55	-0.0014 -		ISM02.4	PPB	7/18/2019 13:01	SMPL
K3767-02	MA52N0	177	0.0287 -		ISM02.4	PPB	7/18/2019 13:03	SMPL
K3767-05	MA52P6	228	0.0413 -		ISM02.4	PPB	7/18/2019 13:05	SMPL
K3767-06	MA52P8	104	0.0107 -		ISM02.4	PPB	7/18/2019 13:07	SMPL
K3767-07	MA52Q4	28	-0.0081 -		ISM02.4	PPB	7/18/2019 13:10	SMPL
K3767-10	MA52S0	61	0.0001 -		ISM02.4	PPB	7/18/2019 13:12	SMPL
K3767-11	MA52S2	14	-0.0115 -		ISM02.4	PPB	7/18/2019 13:14	SMPL
K3767-12	MA52U8	17809	4.3765 -		ISM02.4	PPB	7/18/2019 13:16	SMPL
K3767-12	MA52U8	17836	4.3831 -		ISM02.4	PPB	7/18/2019 13:19	SMPL
K3767-27	MA52S2D	-34	-0.0233 -		ISM02.4	PPB	7/18/2019 13:21	SMPL
K3767-28	MA52S2S	4413	1.0732 -		ISM02.4	PPB	7/18/2019 13:23	SMPL
CCV45	CCV45	20357	5.0048 -		ISM02.4	PPB	7/18/2019 13:26	SMPL
CCB45	CCB45	17	-0.0108 -		ISM02.4	PPB	7/18/2019 13:28	SMPL