

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 07/19/2019

OVEN TEMP1: 95 °C TEMP: 104 °C

DATE & TIME IN: 07/18/2019 17:00 OUT: 07/19/2019 07:25

OVEN TEMP2: 104 °C TEMP: 104 °C

DATE & TIME IN: 07/19/2019 09:30 OUT: 07/19/2019 12:30

Run Number: LB104050

BalanceID: WC SC-3

OvenID: WET OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB104050BL	LB104050BL	98.5013	98.5013	100	98.5013	98.5013	98.5013	0	0
2	K3834-15	TW-3-DEP	92.8078	92.8078	100	92.8429	92.8429	92.8429	0.0351	351
3	K3834-15DUP	TW-3-DEPDUP	95.0723	95.0723	100	95.1077	95.1077	95.1077	0.0354	354
4	K3905-01	18107-NYCT-AVE-Z-TWP-1	85.447	85.447	100	88.7489	88.7489	88.7489	3.3019	33019

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

28104050

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS K3834 WorkList ID : 128722 Department : Wet-Chemistry Date : 07-18-2019 11:53:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2019	Water	K3834-15 -F	TS	Cool 4 deg C	AKRF02		TW-3-DEP	07/17/2019	SM2540 B
07/19/2019	Water	K3905-01 -C	TS	Cool 4 deg C	DIST01		18107-NYCT-AVE-Z-TWP-1	07/18/2019	SM2540 B

Date/Time 07-18-19 16:50
Received by: JS
Relinquished by: CO

Date/Time 07-18-19 17:20
Received by: WJ
Relinquished by: JD