



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB104052

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP76485
Turbidity Calibration std, 0NTU	WP76484
Turbidity Calibration std, 20NTU	WP76480
Turbidity Calibration std, 80NTU	WP76478
Turbidity Calibration std, 40NTU	WP76479
Turbidity Calibration std, 200NTU	WP76477
Turbidity Calibration std, 1NTU	WP76483
Turbidity Calibration std, 10NTU	WP76481
Turbidity Calibration std, 5NTU	WP76482
10 NTU Standard 500 ml	W2297

Intercept: 0.2672 Slope: 0.9077 Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.280	07/18/2019	12:53
2	CAL2	1	1	1.020	07/18/2019	12:57
3	CAL3	5	1	4.550	07/18/2019	12:59
4	CAL4	10	1	9.520	07/18/2019	13:05
5	CAL5	20	1	18.500	07/18/2019	13:12
6	CAL6	40	1	37.500	07/18/2019	13:17
7	CAL7	80	1	71.900	07/18/2019	14:21
8	CAL8	200	1	182.000	07/18/2019	14:26



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB104052

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.550	10.227	07/18/2019	13:22
2	ICB		1	0.238	-0.032	07/18/2019	13:28
3	CCV1	10	1	9.600	10.282	07/18/2019	13:33
4	CCB1		1	0.275	0.009	07/18/2019	13:37
5	LB104052BL		1	0.270	0.003	07/18/2019	13:42
6	K3591-12		1	0.692	0.468	07/18/2019	13:46
7	K3860-02		1	3.800	3.892	07/18/2019	13:55
8	K3882-01		1	9.290	9.940	07/18/2019	13:59
9	K3882-01DUP		1	9.280	9.929	07/18/2019	14:02
10	K3883-01		1	110.000	120.891	07/18/2019	14:06
11	CCV2	10	1	9.610	10.293	07/18/2019	14:10
12	CCB2		1	0.265	-0.002	07/18/2019	14:15

68104052

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tur071819 WorkList ID : 128731 Department : Wet-Chemistry Date : 07-18-2019 12:47:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/30/2019	Water	K3591-12	Turbidity	Cool 4 deg C	CHEM02	QA Office	MDL-MDL-WATER-06-QT3-2C	06/28/2019	SM2130 B
07/19/2019	Water	K3882-01	Turbidity	Cool 4 deg C	APT101	M51	U3MW-5S	07/17/2019	SM2130 B
07/19/2019	Water	K3883-01	Turbidity	Cool 4 deg C	APT101	M51	SGMW-4D	07/17/2019	SM2130 B

Date/Time 07-18-19 12:50
 Received by: AK
 Relinquished by: [Signature]

Date/Time 07-18-19 14:30
 Received by: [Signature]
 Relinquished by: [Signature]