



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB104053

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP76485
TURBIDITY LOD STD, 0.5NTU	WP76486
Turbidity Calibration std, 0NTU	WP76484
Turbidity Calibration std, 20NTU	WP76480
Turbidity Calibration std, 40NTU	WP76479
Turbidity Calibration std, 1NTU	WP76483
Turbidity Calibration std, 10NTU	WP76481
Turbidity Calibration std, 5NTU	WP76482
10 NTU Standard 500 ml	W2297

Intercept: 0.0818

Slope: 0.9326

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.280	07/18/2019	12:53
2	CAL2	1	1	1.020	07/18/2019	12:57
3	CAL3	5	1	4.550	07/18/2019	12:59
4	CAL4	10	1	9.520	07/18/2019	13:05
5	CAL5	20	1	18.500	07/18/2019	13:12
6	CAL6	40	1	37.500	07/18/2019	13:17



Analytical Summary Report

Analysis Method: 180.1
Parameter: Turbidity
Run Number: LB104053

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.550	10.152	07/18/2019	13:28
2	ICB		1	0.238	0.167	07/18/2019	13:28
3	CCV1	10	1	9.580	10.185	07/18/2019	14:24
4	CCB1		1	0.284	0.217	07/18/2019	14:28
5	LB104053BL		1	0.275	0.207	07/18/2019	14:35
6	K3591-09		1	0.628	0.586	07/18/2019	14:41
7	K3860-01		1	3.780	3.965	07/18/2019	14:46
8	K3860-01DUP		1	3.720	3.901	07/18/2019	14:50
9	CCV2	10	1	9.520	10.120	07/18/2019	14:55
10	CCB2		1	0.276	0.208	07/18/2019	14:59

WORKLIST(Hardcopy Internal Chain)

69104053

WorkList Name : tur071719-2 WorkList ID : 128665 Department : Wet-Chemistry Date : 07-17-2019 15:24:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/30/2019	Water	K3591-09	Turbidity	Cool 4 deg C	CHEM02	QA Office	MDL-MDL-WATER-03-QT3-2C	06/28/2019	180.1

Date/Time 07-18-19
 Received by: N/A
 Relinquished by: _____

Date/Time 07-18-19
 Received by: N/A
 Relinquished by: _____