


LB104059

7471B

INSTRUMENT ID : CVI

Linear



A= 0.0000e+000

B= 1.7611e-004

C= -2.5249e-002

Rho= 0.9999900

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	231	0.000	231				
0.2	0.200	0.203	0.003	1297	0.0 %	1297				
2.5	2.500	2.497	-0.003	14322	0.0 %	14322				
5.0	5.000	4.966	-0.034	28343	0.0 %	28343				
7.5	7.500	7.503	0.003	42745	0.0 %	42745				
10.0	10.000	10.015	0.015	57012	0.0 %	57012				

LB104059 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Jul 2019 15:33:23

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0-0	231	-	0.00007471B	PPB		18 Jul 2019 15:33:31	S	Std
0.2 - 1	S0-2	1297	-	0.20007471B	PPB		18 Jul 2019 15:35:49	S	Std
2.5 - 1	S2-5	14322	-	2.50007471B	PPB		18 Jul 2019 15:38:07	S	Std
5.0 - 1	S5-0	28343	-	5.00007471B	PPB		18 Jul 2019 15:40:24	S	Std
7.5 - 1	S7-5	42745	-	7.50007471B	PPB		18 Jul 2019 15:42:42	S	Std
10.0 - 1	S10-0	57012	-	10.00007471B	PPB		18 Jul 2019 15:44:58	S	Std
ICV38 - 1	ICV38	22779	3.9864	-7471B	PPB		18 Jul 2019 15:48:56	U	SMPL
ICB38 - 1	ICB38	163	0.0035	-7471B	PPB		18 Jul 2019 15:51:11	U	SMPL
CCV46 - 1	CCV46	28230	4.9464	-7471B	PPB		18 Jul 2019 15:53:28	U	SMPL
CCB46 - 1	CCB46	162	0.0033	-7471B	PPB		18 Jul 2019 15:55:44	U	SMPL
CRA - 1	CRA	1287	0.2014	-7471B	PPB		18 Jul 2019 15:58:00	U	SMPL
HighStd - 1	HighStd	56024	9.8413	-7471B	PPB		18 Jul 2019 16:00:16	U	SMPL
ChkStd - 1	ChkStd	39552	6.9404	-7471B	PPB		18 Jul 2019 16:02:32	U	SMPL
PB121508BL - 1	PBS	83	-0.0106	-7471B	PPB		18 Jul 2019 16:04:49	U	SMPL
PB121508BS - 1	LCSS	20202	3.5326	-7471B	PPB		18 Jul 2019 16:07:07	U	SMPL
K3835-01 - 1	SB-1(0-2)	17810	3.1113	-7471B	PPB		18 Jul 2019 16:09:23	U	SMPL
K3835-01DUP - 1	SB-1(0-2)DUP	17677	3.0879	-7471B	PPB		18 Jul 2019 16:11:41	U	SMPL
K3835-01MS - 1	SB-1(0-2)MS	36244	6.3578	-7471B	PPB		18 Jul 2019 16:13:58	U	SMPL
K3835-01MSD - 1	SB-1(0-2)MSD	35839	6.2865	-7471B	PPB		18 Jul 2019 16:16:16	U	SMPL
K3835-02 - 1	SB-1(7-9)	824	0.1199	-7471B	PPB		18 Jul 2019 16:18:33	U	SMPL
CCV47 - 1	CCV47	28041	4.9131	-7471B	PPB		18 Jul 2019 16:20:51	U	SMPL
CCB47 - 1	CCB47	156	0.0022	-7471B	PPB		18 Jul 2019 16:23:07	U	SMPL
K3835-03 - 1	SB-2(0-2)	3032	0.5087	-7471B	PPB		18 Jul 2019 16:25:24	U	SMPL
K3835-04 - 1	SB-2(10-12)	587	0.0781	-7471B	PPB		18 Jul 2019 16:27:39	U	SMPL
K3835-05 - 1	SB-3(0-2)	37318	6.5469	-7471B	PPB		18 Jul 2019 16:29:55	U	SMPL
K3835-06 - 1	SB-3(6-8)	412	0.0473	-7471B	PPB		18 Jul 2019 16:32:11	U	SMPL
K3835-07 - 1	SB-4(0-2)	1060	0.1614	-7471B	PPB		18 Jul 2019 16:34:30	U	SMPL
K3835-08 - 1	SB-4(7-9)	569	0.0750	-7471B	PPB		18 Jul 2019 16:36:51	U	SMPL
K3835-09 - 1	SB-5(0-2)	884	0.1304	-7471B	PPB		18 Jul 2019 16:39:09	U	SMPL
K3835-10 - 1	SB-5(7-9)	333	0.0334	-7471B	PPB		18 Jul 2019 16:41:27	U	SMPL
K3835-23 - 1	SB-12(0-2)	8381	1.4508	-7471B	PPB		18 Jul 2019 16:43:44	U	SMPL
K3835-24 - 1	SB-12(6-8)	594	0.0794	-7471B	PPB		18 Jul 2019 16:46:00	U	SMPL
CCV48 - 1	CCV48	28012	4.9080	-7471B	PPB		18 Jul 2019 16:48:16	U	SMPL
CCB48 - 1	CCB48	228	0.0149	-7471B	PPB		18 Jul 2019 16:50:32	U	SMPL
K3835-25 - 1	SB-9(0-2)	1038	0.1576	-7471B	PPB		18 Jul 2019 16:52:48	U	SMPL
K3835-26 - 1	SB-9(5-7)	1399	0.2211	-7471B	PPB		18 Jul 2019 16:55:04	U	SMPL
K3835-27 - 1	SB-10(0-2)	5370	0.9205	-7471B	PPB		18 Jul 2019 16:57:20	U	SMPL
K3835-28 - 1	SB-10(8-10)	486	0.0603	-7471B	PPB		18 Jul 2019 16:59:36	U	SMPL
K3835-29 - 1	SB-11(0-2)	26218	4.5921	-7471B	PPB		18 Jul 2019 17:01:53	U	SMPL
K3835-30 - 1	SB-11(8-10)	253	0.0193	-7471B	PPB		18 Jul 2019 17:04:09	U	SMPL
K3835-31 - 1	SB-7(0-2)	33154	5.8136	-7471B	PPB		18 Jul 2019 17:06:28	U	SMPL
K3835-32 - 1	SB-7(13-15)	385	0.0426	-7471B	PPB		18 Jul 2019 17:08:45	U	SMPL
PB121509BL - 1	PBS	158	0.0026	-7471B	PPB		18 Jul 2019 17:11:03	U	SMPL
PB121509BS - 1	LCSS	20233	3.5380	-7471B	PPB		18 Jul 2019 17:13:19	U	SMPL
CCV49 - 1	CCV49	28296	4.9580	-7471B	PPB		18 Jul 2019 17:15:35	U	SMPL
CCB49 - 1	CCB49	224	0.0142	-7471B	PPB		18 Jul 2019 17:17:51	U	SMPL
K3615-01 - 1	HR-06-071719-A	4662	0.7958	-7471B	PPB		18 Jul 2019 17:20:09	U	SMPL
K3615-03 - 1	HR-06-071719-B	4253	0.7238	-7471B	PPB		18 Jul 2019 17:22:25	U	SMPL
K3615-05 - 1	HR-06-071719-C	3989	0.6773	-7471B	PPB		18 Jul 2019 17:24:41	U	SMPL
K3618-05 - 1	HD-01-071719	1483	0.2359	-7471B	PPB		18 Jul 2019 17:26:57	U	SMPL
K3835-43 - 1	SB-6(0-2)	9971	1.7308	-7471B	PPB		18 Jul 2019 17:29:13	U	SMPL
K3835-44 - 1	SB-6(8-10)	597	0.0799	-7471B	PPB		18 Jul 2019 17:31:30	U	SMPL
K3835-45 - 1	SB-8(0-2)	6592	1.1357	-7471B	PPB		18 Jul 2019 17:33:47	U	SMPL
K3835-46 - 1	SB-8(4-6)	971	0.1458	-7471B	PPB		18 Jul 2019 17:36:05	U	SMPL
K3838-01 - 1	WARINANCO-CITY-SEWER-INSTALL	1014	0.1533	-7471B	PPB		18 Jul 2019 17:38:23	U	SMPL
K3838-01DUP - 1	WARINANCO-CITY-SEWER-INSTALLDUP	1012	0.1530	-7471B	PPB		18 Jul 2019 17:40:39	U	SMPL
CCV50 - 1	CCV50	28560	5.0045	-7471B	PPB		18 Jul 2019 17:42:55	U	SMPL
CCB50 - 1	CCB50	249	0.0186	-7471B	PPB		18 Jul 2019 17:45:12	U	SMPL
K3838-01MS - 1	WARINANCO-CITY-SEWER-INSTALLMS	21641	3.7860	-7471B	PPB		18 Jul 2019 17:47:28	U	SMPL
K3838-01MSD - 1	WARINANCO-CITY-SEWER-INSTALLMSD	21386	3.7411	-7471B	PPB		18 Jul 2019 17:49:45	U	SMPL
K3840-01 - 1	FK-SF-02-005-A	6582	1.1339	-7471B	PPB		18 Jul 2019 17:52:03	U	SMPL
K3840-03 - 1	FK-SF-02-005-B	5201	0.8907	-7471B	PPB		18 Jul 2019 17:54:19	U	SMPL
K3852-01 - 1	COMP-1_071619	5043	0.8629	-7471B	PPB		18 Jul 2019 17:56:35	U	SMPL
K3852-02 - 1	COMP-2_071619	8665	1.5008	-7471B	PPB		18 Jul 2019 17:58:52	U	SMPL
K3852-03 - 1	COMP-3_071619	24367	4.2661	-7471B	PPB		18 Jul 2019 18:01:10	U	SMPL
K3866-01 - 1	RT-1488	1304	0.2044	-7471B	PPB		18 Jul 2019 18:03:28	U	SMPL
K3868-05 - 1	TP-1	3357	0.5660	-7471B	PPB		18 Jul 2019 18:05:48	U	SMPL
K3868-07 - 1	TP-2	5648	0.9694	-7471B	PPB		18 Jul 2019 18:08:05	U	SMPL
CCV51 - 1	CCV51	28411	4.9783	-7471B	PPB		18 Jul 2019 18:10:21	U	SMPL
CCB51 - 1	CCB51	140	-0.0006	-7471B	PPB		18 Jul 2019 18:12:38	U	SMPL
K3868-09 - 1	TP-3	6489	1.1175	-7471B	PPB		18 Jul 2019 18:14:54	U	SMPL
K3868-11 - 1	TP-4	4674	0.7979	-7471B	PPB		18 Jul 2019 18:17:11	U	SMPL
K3835-01LX5 - 1		3472	0.5862	-7471B	PPB		18 Jul 2019 18:19:27	U	SMPL
K3835-01A - 1		36494	6.4018	-7471B	PPB		18 Jul 2019 18:21:45	U	SMPL
K3838-01LX5 - 1		260	0.0205	-7471B	PPB		18 Jul 2019 18:24:02	U	SMPL
K3838-01A - 1		21610	3.7806	-7471B	PPB		18 Jul 2019 18:26:20	U	SMPL
CCV52 - 1	CCV52	28305	4.9596	-7471B	PPB		18 Jul 2019 18:28:37	U	SMPL
CCB52 - 1	CCB52	135	-0.0015	-7471B	PPB		18 Jul 2019 18:30:55	U	SMPL