



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104063

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/18/2019	18:00
2	CAL2	1	Water	NA	NA	21.2	7.01	07/18/2019	18:01
3	CAL3	1	Water	NA	NA	21.2	10.03	07/18/2019	18:02
4	ICV	1	Water	NA	NA	21.2	6.99	07/18/2019	18:04
5	CCV1	1	Water	NA	NA	21.2	2.02	07/18/2019	18:06
6	K3881-02	1	Solid	20.03	20	22.6	7.82	07/18/2019	18:07
7	K3886-02	1	Solid	20.01	20	24.2	5.40	07/18/2019	18:09
8	K3887-02	1	Solid	20.04	20	23.3	8.32	07/18/2019	18:10
9	K3887-04	1	Solid	20.02	20	23.2	8.39	07/18/2019	18:11
10	K3887-06	1	Solid	20.03	20	22.7	7.90	07/18/2019	18:13
11	K3887-08	1	Solid	20.04	20	22.6	7.94	07/18/2019	18:15
12	K3900-02	1	Solid	20.04	20	24.1	6.54	07/18/2019	18:16
13	K3903-07	1	Solid	20.03	20	23.3	8.29	07/18/2019	18:19
14	K3903-08	1	Solid	20.05	20	22.9	8.60	07/18/2019	18:20
15	K3903-08DUP	1	Solid	20.04	20	22.8	8.61	07/18/2019	18:22
16	CCV2	1	Water	NA	NA	21.2	12.02	07/18/2019	18:25

WORKLIST(Hardcopy Internal Chain)

VB 104063

WorkList Name : CORROSIONITY K3903 WorkList ID : 128744 Department : Wet-Chemistry Date : 07-18-2019 16:46:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/17/2019	Solid	K3835-49	Corrosivity	Cool 4 deg C	AKRF02		SB-8(0-2)	07/17/2019	9045D
07/17/2019	Solid	K3881-02	Corrosivity	Cool 4 deg C	DIST01	M61	SUP-SB-C-1-071719	07/17/2019	9045D
07/18/2019	Solid	K3886-02	Corrosivity	Cool 4 deg C	PSEG01	M61	DEBRIS	07/18/2019	9045D
07/18/2019	Solid	K3887-02	Corrosivity	Cool 4 deg C	TULL02	M62	AOC-7(1)	07/18/2019	9045D
07/18/2019	Solid	K3887-04	Corrosivity	Cool 4 deg C	TULL02	M62	AOC-7(2)	07/18/2019	9045D
07/18/2019	Solid	K3887-06	Corrosivity	Cool 4 deg C	TULL02	M62	AOC-7(3)	07/18/2019	9045D
07/18/2019	Solid	K3887-08	Corrosivity	Cool 4 deg C	TULL02	M62	AOC-7(4)	07/18/2019	9045D
07/18/2019	Solid	K3900-02	Corrosivity	Cool 4 deg C	PSEG01	M61	SOIL-COMP	07/18/2019	9045D
07/18/2019	Solid	K3903-07	Corrosivity	Cool 4 deg C	DIST01		180107-NYCT-AVE-Z-COMP-	07/18/2019	9045D
07/18/2019	Solid	K3903-08	Corrosivity	Cool 4 deg C	DIST01		180107-NYCT-AVE-Z-COMP-	07/18/2019	9045D

Date/Time 07-18-19 16:50
 Received by: JA
 Relinquished by: CA

Date/Time 07-18-19 18:35
 Received by: CO
 Relinquished by: JA