



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB104079

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP76533
Turbidity Calibration std, 0NTU	WP76525
Turbidity Calibration std, 20NTU	WP76528
Turbidity Calibration std, 80NTU	WP76531
Turbidity Calibration std, 40NTU	WP76527
Turbidity Calibration std, 200NTU	WP76532
Turbidity Calibration std, 1NTU	WP76534
Turbidity Calibration std, 10NTU	WP76529
Turbidity Calibration std, 5NTU	WP76530
10 NTU Standard 500 ml	W2297

Intercept: 0.0921

Slope: 0.9323

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.314	07/19/2019	10:45
2	CAL2	1	1	1.020	07/19/2019	10:49
3	CAL3	5	1	4.550	07/19/2019	10:55
4	CAL4	10	1	9.520	07/19/2019	11:01
5	CAL5	20	1	18.500	07/19/2019	11:06
6	CAL6	40	1	37.500	07/19/2019	11:12
7	CAL7	80	1	72.100	07/19/2019	11:56
8	CAL8	200	1	182.200	07/19/2019	11:59



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB104079

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.450	10.037	07/19/2019	11:16
2	ICB		1	0.296	0.219	07/19/2019	11:23
3	CCV1	10	1	9.560	10.155	07/19/2019	11:27
4	CCB1		1	0.302	0.225	07/19/2019	11:30
5	LB104079BL		1	0.350	0.277	07/19/2019	11:34
6	K3908-01		1	145.000	155.431	07/19/2019	11:39
7	K3908-01DUP		1	144.000	154.358	07/19/2019	11:43
8	CCV2	10	1	9.510	10.102	07/19/2019	11:48
9	CCB2		1	0.328	0.253	07/19/2019	11:53

WORKLIST(Hardcopy Internal Chain)

LB104079

WorkList Name : tur071919 WorkList ID : 128778 Department : Wet-Chemistry Date : 07-19-2019 10:06:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/20/2019	Water	K3908-01	Turbidity	Cool 4 deg C	APT101	M52	PWU-3	07/18/2019	SM2130 B

Date/Time 07-19-19 10:10
 Received by: AO
 Relinquished by: JP

Date/Time 07-19-19 12:05
 Received by: JP
 Relinquished by: AO