



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB104087

Analyst By : JIGNESH

Supervisor Review By : apatel

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/18/2019	17:15
2	CAL2	1	Water	NA	NA	20.2	7.01	07/18/2019	17:16
3	CAL3	1	Water	NA	NA	20.2	10.03	07/18/2019	17:17
4	ICV	1	Water	NA	NA	20.2	6.99	07/18/2019	17:18
5	CCV1	1	Water	NA	NA	20.2	2.02	07/18/2019	17:20
6	K2646-09	1	Water	NA	NA	23.7	6.16	07/18/2019	17:22
7	K2646-10	1	Water	NA	NA	24.1	7.86	07/18/2019	17:24
8	K2646-16	1	Water	NA	NA	23.2	4.91	07/18/2019	17:25
9	K2646-17	1	Water	NA	NA	23.1	6.76	07/18/2019	17:26
10	K2646-18	1	Water	NA	NA	23.1	6.91	07/18/2019	17:29
11	K2646-19	1	Water	NA	NA	23.6	5.94	07/18/2019	17:30
12	K2646-20	1	Water	NA	NA	24.2	6.21	07/18/2019	17:32
13	K2646-22	1	Water	NA	NA	23.0	6.52	07/18/2019	17:34
14	K2646-23	1	Water	NA	NA	23.0	5.80	07/18/2019	17:35
15	K3901-01	1	Water	NA	NA	23.5	7.09	07/18/2019	17:36
16	CCV2	1	Water	NA	NA	20.1	12.01	07/18/2019	17:38
17	K3901-01DUP	1	Water	NA	NA	23.6	7.10	07/18/2019	17:40
18	CCV3	1	Water	NA	NA	20.2	2.02	07/18/2019	17:41

LB 10408f

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W K2646 WorkList ID : 128793 Department : Wet-Chemistry Date : 07-18-2019 17:00:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/18/2019	Water	K2646-09 ✓	pH	Cool 4 deg C	PSEG01	M62	30341	07/18/2019	9040C
07/18/2019	Water	K2646-10 ✓	pH	Cool 4 deg C	PSEG01	M62	30369	07/18/2019	9040C
07/18/2019	Water	K2646-16 ✓	pH	Cool 4 deg C	PSEG01	M62	30297	07/18/2019	9040C
07/18/2019	Water	K2646-17 ✓	pH	Cool 4 deg C	PSEG01	M62	30298	07/18/2019	9040C
07/18/2019	Water	K2646-18 ✓	pH	Cool 4 deg C	PSEG01	M62	30299	07/18/2019	9040C
07/18/2019	Water	K2646-19 ✓	pH	Cool 4 deg C	PSEG01	M62	30301	07/18/2019	9040C
07/18/2019	Water	K2646-20 ✓	pH	Cool 4 deg C	PSEG01	M62	COMP-4	07/18/2019	9040C
07/18/2019	Water	K2646-22 ✓	pH	Cool 4 deg C	PSEG01	M62	30355	07/18/2019	9040C
07/18/2019	Water	K2646-23 ✓	pH	Cool 4 deg C	PSEG01	M62	30356	07/18/2019	9040C
07/18/2019	Water	K3901-01 ✓	pH	Cool 4 deg C	PSEG01	M61	CONTAINMENT-WATER	07/18/2019	9040C

Date/Time 07-18-19 17:05
 Received by: JG
 Relinquished by: JG

Date/Time 07-18-19 18:40
 Received by: JG
 Relinquished by: JG