



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104088

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/18/2019	17:15
2	CAL2	1	Water	NA	NA	20.2	7.01	07/18/2019	17:16
3	CAL3	1	Water	NA	NA	20.2	10.03	07/18/2019	17:17
4	ICV	1	Water	NA	NA	20.2	6.99	07/18/2019	17:18
5	CCV1	1	Water	NA	NA	20.2	2.02	07/18/2019	17:20
6	K3901-02	1	Water	NA	NA	23.5	7.09	07/18/2019	17:36
7	K3901-02DUP	1	Water	NA	NA	23.6	7.10	07/18/2019	17:40
8	CCV2	1	Water	NA	NA	20.2	12.01	07/18/2019	17:41

13104088

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K3901 W WorkList ID : 128794 Department : Wet-Chemistry Date : 07-18-2019 17:00:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/18/2019	Water	K3901-02	☒ Corrosivity	Cool 4 deg C	PSEG01	M61	CONTAINMENT-WATER	07/18/2019	9040C

Date/Time 07-18-19 17:05
Received by: JJ
Relinquished by: JJ

Date/Time 07-18-19 18:40
Received by: JJ
Relinquished by: JJ