


LB104097

7470A

INSTRUMENT ID : C01



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.012	0.012	73	0.000	73				
0.2	0.200	0.175	-0.025	736	0.0 %	736				
2.5	2.500	2.390	-0.110	9762	0.0 %	9762				
5.0	5.000	5.143	0.143	20979	0.0 %	20979				
7.5	7.500	7.623	0.123	31087	0.0 %	31087				
10.0	10.000	9.857	-0.143	40190	0.0 %	40190				

LB104097 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 19 Jul 2019 14:55:51			
Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type
0.0 - 1	50-0	73	-	0.00007470A	PPB	19 Jul 2019 15:04:40	S Std
0.2 - 1	50-2	736	-	0.20007470A	PPB	19 Jul 2019 15:06:48	S Std
2.5 - 1	52-5	9762	-	2.50007470A	PPB	19 Jul 2019 15:08:57	S Std
5.0 - 1	55-0	20979	-	5.00007470A	PPB	19 Jul 2019 15:11:06	S Std
7.5 - 1	57-5	31087	-	7.50007470A	PPB	19 Jul 2019 15:13:13	S Std
10.0 - 1	510-0	40190	-	10.00007470A	PPB	19 Jul 2019 15:27:55	S Std
ICV41 - 1	ICV41	15911	3.8989	-7470A	PPB	19 Jul 2019 15:30:52	U SMP
ICB41 - 1	ICB41	-14	-0.0093	-7470A	PPB	19 Jul 2019 15:33:00	U SMP
CCV59 - 1	CCV59	20923	5.1289	-7470A	PPB	19 Jul 2019 15:35:06	U SMP
CCB59 - 1	CCB59	5	-0.0046	-7470A	PPB	19 Jul 2019 15:37:13	U SMP
CRA - 1	CRA	714	0.1694	-7470A	PPB	19 Jul 2019 15:39:19	U SMP
HighStd - 1	HighStd	41320	10.1346	-7470A	PPB	19 Jul 2019 15:41:26	U SMP
ChkStd - 1	ChkStd	28466	6.9801	-7470A	PPB	19 Jul 2019 15:43:33	U SMP
PB121549BL - 1	PBW	-9	-0.0081	-7470A	PPB	19 Jul 2019 15:45:41	U SMP
PB121549BS - 1	LCSW	16303	3.9951	-7470A	PPB	19 Jul 2019 15:47:49	U SMP
K3615-02 - 1	HR-06-071719-A	8	-0.0039	-7470A	PPB	19 Jul 2019 15:49:56	U SMP
K3615-02DUP - 1	HR-06-071719-ADUP	37	0.0032	-7470A	PPB	19 Jul 2019 15:52:04	U SMP
K3615-02MS - 1	HR-06-071719-AMS	15668	3.8393	-7470A	PPB	19 Jul 2019 15:54:12	U SMP
K3615-02MSD - 1	HR-06-071719-AMSD	16381	4.0143	-7470A	PPB	19 Jul 2019 15:56:21	U SMP
K3615-04 - 1	HR-06-071719-B	153	0.0317	-7470A	PPB	19 Jul 2019 15:58:28	U SMP
CCV60 - 1	CCV60	21206	5.1984	-7470A	PPB	19 Jul 2019 16:00:34	U SMP
CCB60 - 1	CCB60	2	-0.0054	-7470A	PPB	19 Jul 2019 16:02:41	U SMP
K3615-06 - 1	HR-06-071719-C	83	0.0145	-7470A	PPB	19 Jul 2019 16:04:48	U SMP
K3618-06 - 1	HD-01-071719	46	0.0054	-7470A	PPB	19 Jul 2019 16:06:56	U SMP
K3866-02 - 1	RT-1488	71	0.0116	-7470A	PPB	19 Jul 2019 16:09:03	U SMP
K3868-06 - 1	TP-1	97	0.0179	-7470A	PPB	19 Jul 2019 16:11:11	U SMP
K3868-08 - 1	TP-2	62	0.0094	-7470A	PPB	19 Jul 2019 16:13:18	U SMP
K3868-10 - 1	TP-3	121	0.0238	-7470A	PPB	19 Jul 2019 16:15:26	U SMP
K3868-12 - 1	TP-4	139	0.0282	-7470A	PPB	19 Jul 2019 16:17:34	U SMP
PB121550BL - 1	PBW	60	0.0089	-7470A	PPB	19 Jul 2019 16:19:42	U SMP
PB121550BS - 1	LCSW	16251	3.9824	-7470A	PPB	19 Jul 2019 16:21:51	U SMP
K3838-04 - 1	WARINANCO-CITY-SEWER-INSTALL	45	0.0052	-7470A	PPB	19 Jul 2019 16:23:58	U SMP
CCV61 - 1	CCV61	21358	5.2357	-7470A	PPB	19 Jul 2019 16:26:04	U SMP
CCB61 - 1	CCB61	9	-0.0037	-7470A	PPB	19 Jul 2019 16:28:11	U SMP
K3857-02 - 1	36916	41	0.0042	-7470A	PPB	19 Jul 2019 16:30:18	U SMP
K3857-02DUP - 1	36916DUP	102	0.0192	-7470A	PPB	19 Jul 2019 16:32:25	U SMP
K3857-02MS - 1	36916MS	15337	3.7581	-7470A	PPB	19 Jul 2019 16:34:32	U SMP
K3857-02MSD - 1	36916MSD	15895	3.8950	-7470A	PPB	19 Jul 2019 16:36:39	U SMP
PB121551BL - 1	PBW	23	-0.0002	-7470A	PPB	19 Jul 2019 16:38:46	U SMP
PB121551BS - 1	LCSW	16472	4.0366	-7470A	PPB	19 Jul 2019 16:40:54	U SMP
K3881-02 - 1	SUP-SB-C-1-071719	575	0.1352	-7470A	PPB	19 Jul 2019 16:43:01	U SMP
K3886-02 - 1	DEBRIS	528	0.1237	-7470A	PPB	19 Jul 2019 16:45:10	U SMP
K3887-02 - 1	AOC-7(1)	493	0.1151	-7470A	PPB	19 Jul 2019 16:47:18	U SMP
K3887-04 - 1	AOC-7(2)	287	0.0646	-7470A	PPB	19 Jul 2019 16:49:25	U SMP
CCV62 - 1	CCV62	19966	4.8941	-7470A	PPB	19 Jul 2019 16:51:32	U SMP
CCB62 - 1	CCB62	10	-0.0034	-7470A	PPB	19 Jul 2019 16:53:43	U SMP
K3887-06 - 1	AOC-7(3)	919	0.2197	-7470A	PPB	19 Jul 2019 16:55:50	U SMP
K3887-08 - 1	AOC-7(4)	764	0.1816	-7470A	PPB	19 Jul 2019 16:57:57	U SMP
K3903-07 - 1	180107-NYCT-AVE-Z-COMP-1-2	650	0.1537	-7470A	PPB	19 Jul 2019 17:00:04	U SMP
K3903-08 - 1	180107-NYCT-AVE-Z-COMP-3-4	511	0.1195	-7470A	PPB	19 Jul 2019 17:02:11	U SMP
K3903-08DUP - 1	180107-NYCT-AVE-Z-COMP-3-4DUP	556	0.1306	-7470A	PPB	19 Jul 2019 17:04:18	U SMP
K3903-08MS - 1	180107-NYCT-AVE-Z-COMP-3-4MS	15805	3.8729	-7470A	PPB	19 Jul 2019 17:06:27	U SMP
K3903-08MSD - 1	180107-NYCT-AVE-Z-COMP-3-4MSD	16614	4.0714	-7470A	PPB	19 Jul 2019 17:08:35	U SMP
PB121552BL - 1	PBW	17	-0.0017	-7470A	PPB	19 Jul 2019 17:10:44	U SMP
PB121552BS - 1	LCSW	16298	3.9939	-7470A	PPB	19 Jul 2019 17:12:52	U SMP
K3834-01 - 1	TW-3	38	0.0035	-7470A	PPB	19 Jul 2019 17:15:03	U SMP
CCV63 - 1	CCV63	20297	4.9753	-7470A	PPB	19 Jul 2019 17:17:10	U SMP
CCB63 - 1	CCB63	23	-0.0002	-7470A	PPB	19 Jul 2019 17:19:17	U SMP
K3834-01DUP - 1	TW-3DUP	65	0.0101	-7470A	PPB	19 Jul 2019 17:21:24	U SMP
K3834-01MS - 1	TW-3MS	16016	3.9247	-7470A	PPB	19 Jul 2019 17:23:31	U SMP
K3834-01MSD - 1	TW-3MSD	16792	4.1151	-7470A	PPB	19 Jul 2019 17:25:37	U SMP
K3834-02 - 1	TW-4	1495	0.3610	-7470A	PPB	19 Jul 2019 17:27:44	U SMP
K3834-04 - 1	TW-3	52	0.0069	-7470A	PPB	19 Jul 2019 17:29:52	U SMP
K3834-05 - 1	TW-4	55	0.0076	-7470A	PPB	19 Jul 2019 17:31:59	U SMP
K3834-08 - 1	TW-5	63	0.0096	-7470A	PPB	19 Jul 2019 17:34:07	U SMP
K3834-09 - 1	TW-9	172	0.0363	-7470A	PPB	19 Jul 2019 17:36:15	U SMP
K3834-11 - 1	TW-5	35	0.0027	-7470A	PPB	19 Jul 2019 17:38:26	U SMP
K3834-12 - 1	TW-9	49	0.0062	-7470A	PPB	19 Jul 2019 17:40:33	U SMP
CCV64 - 1	CCV64	19643	4.8148	-7470A	PPB	19 Jul 2019 17:42:43	U SMP
CCB64 - 1	CCB64	8	-0.0039	-7470A	PPB	19 Jul 2019 17:44:52	U SMP
K3834-13 - 1	TW-8	94	0.0172	-7470A	PPB	19 Jul 2019 17:47:01	U SMP
K3834-14 - 1	TW-7	126	0.0251	-7470A	PPB	19 Jul 2019 17:49:09	U SMP
K3834-16 - 1	TW-6	122	0.0241	-7470A	PPB	19 Jul 2019 17:51:19	U SMP
K3834-18 - 1	TW-8	43	0.0047	-7470A	PPB	19 Jul 2019 17:53:29	U SMP
K3834-19 - 1	TW-7	47	0.0057	-7470A	PPB	19 Jul 2019 17:55:37	U SMP
K3834-20 - 1	TW-6	59	0.0086	-7470A	PPB	19 Jul 2019 17:57:51	U SMP
K3857-01 - 1	36916	50	0.0064	-7470A	PPB	19 Jul 2019 18:00:00	U SMP
K3884-13 - 1	FL-0629	330	0.0751	-7470A	PPB	19 Jul 2019 18:02:09	U SMP
K3901-01 - 1	CONTAINMENT-WATER	280	0.0629	-7470A	PPB	19 Jul 2019 18:04:21	U SMP
PB121549TB - 1	PB121462TB	55	0.0076	-7470A	PPB	19 Jul 2019 18:06:30	U SMP
CCV65 - 1	CCV65	20391	4.9984	-7470A	PPB	19 Jul 2019 18:08:41	U SMP
CCB65 - 1	CCB65	-24	-0.0118	-7470A	PPB	19 Jul 2019 18:10:49	U SMP
PB121550TB - 1	PB121451TB	59	0.0086	-7470A	PPB	19 Jul 2019 18:12:57	U SMP
PB121551TB - 1	PB121505TB	30	0.0015	-7470A	PPB	19 Jul 2019 18:15:11	U SMP
K3615-02LX5 - 1		32	0.0020	-7470A	PPB	19 Jul 2019 18:23:18	U SMP
K3615-02A - 1		15888	3.8933	-7470A	PPB	19 Jul 2019 18:25:26	U SMP
K3857-02LX5 - 1		11	-0.0032	-7470A	PPB	19 Jul 2019 18:27:33	U SMP
K3857-02A - 1		15237	3.7335	-7470A	PPB	19 Jul 2019 18:29:40	U SMP
K3903-08LX5 - 1		116	0.0226	-7470A	PPB	19 Jul 2019 18:31:49	U SMP
K3903-08A - 1		15958	3.9105	-7470A	PPB	19 Jul 2019 18:33:57	U SMP
K3834-01LX5 - 1		-3	-0.0066	-7470A	PPB	19 Jul 2019 18:36:06	U SMP
K3834-01A - 1		16200	3.9698	-7470A	PPB	19 Jul 2019 18:38:14	U SMP
CCV66 - 1	CCV66	20717	5.0784	-7470A	PPB	19 Jul 2019 18:40:21	U SMP

LB104097 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 19 Jul 2019 14:55:51			
Sample ID	Extended ID	μ Abs	Conc	Strd Conc	Method	Units	Date Type Type
CCB66 - 1	CCB66	38	0.0035		7470A	PPB	19 Jul 2019 18:42:28 U SMPL