



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 7/22/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:26
In Date: 07/19/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 07/20/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104102

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3617-01	NB-07-071919-A	1	1.13	9.75	8.91	90.3
K3617-03	NB-07-071919-B	2	1.16	9.88	8.86	88.3
K3621-05	OK-02-071919	3	1.15	9.95	9.6	96.0
K3622-09	OR-02-071919-A	4	1.11	9.86	8.9	89.0
K3622-11	OR-02-071919-B	5	1.13	9.86	8.95	89.6
K3891-07	P001-SW03-01	6	1.00	2.00	2.00	100.0
K3896-06	TRAN-134	7	1.00	2.00	2.00	100.0
K3896-07	TRAN-135	8	1.00	2.00	2.00	100.0
K3896-08	TRAN-105	9	1.00	2.00	2.00	100.0
K3896-09	TRAN-104	10	1.00	2.00	2.00	100.0
K3903-01	18107-NYCT-AVE-Z-TP-1	11	1.13	6.87	6.09	86.4
K3903-02	18107-NYCT-AVE-Z-TP-2	12	1.15	6.53	5.62	83.1
K3903-03	180107-NYCT-AVE-Z-TP-3	13	1.13	4.66	4.32	90.4
K3903-04	180107-NYCT-AVE-Z-TP-4	14	1.13	3.25	3.05	90.6
K3911-01	HJDC-G1	15	1.19	9.79	8.59	86.0
K3911-02	HJDC-G2	16	1.11	9.71	8.67	87.9
K3911-03	HJDC-COMP1	17	1.15	9.94	8.98	89.1
K3914-01	S10-4-B	18	1.15	9.77	8.36	83.6
K3914-02	S10-4-S	19	1.12	9.93	8.18	80.1
K3914-03	S15-3-B	20	1.11	9.58	8.65	89.0
K3914-04	S15-3-S	21	1.19	9.68	8.02	80.4
K3916-01	GIS-BUILDING-1	22	1.11	9.98	8.49	83.2
K3916-02	GIS-BUILDING-2	23	1.11	9.97	8.47	83.1
K3916-03	GIS-BUILDING-3	24	1.19	9.98	8.69	85.3
K3917-05	SVOC-GPC-BLANK	25	1.00	2.00	2.00	100.0
K3917-06	PEST-GPC-BLANK	26	1.00	2.00	2.00	100.0
K3917-07	PEST-GPC-BLANK-SPIKE	27	1.00	2.00	2.00	100.0
K3917-08	PCB-GPC-BLANK	28	1.00	2.00	2.00	100.0
K3917-09	PCB-GPC-BLANK-SPIKE	29	1.00	2.00	2.00	100.0
K3917-10	SVOC-GPC2-BLANK	30	1.00	2.00	2.00	100.0
K3917-11	PEST-GPC2-BLANK	31	1.00	2.00	2.00	100.0
K3917-12	PEST-GPC2-BLANK-SPIKE	32	1.00	2.00	2.00	100.0
K3917-13	PCB-GPC2-BLANK	33	1.00	2.00	2.00	100.0
K3917-14	PCB-GPC2-BLANK -SPIKE	34	1.00	2.00	2.00	100.0
K3917-15	FLORISIL-CHECK	35	1.00	2.00	2.00	100.0
K3918-01	S3-03-S	36	1.14	9.63	7.85	79.0
K3918-02	S3-03-B	37	1.15	9.96	9.00	89.1
K3918-03	S3-02-S	38	1.11	9.58	9.18	95.3
K3918-04	S3-02-B	39	1.19	9.77	9.04	91.5



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 7/22/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:26
In Date: 07/19/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 07/20/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104102

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3918-05	S3-01-S	40	1.13	9.9	8.49	83.9
K3918-06	S3-01-B	41	1.13	9.71	8.27	83.2
K3918-07	DUP-1	42	1.13	9.86	9.3	93.6
K3927-01	01-A-01-B-01-C	43	1.00	2.00	2.00	100.0
K3927-02	02-A-02-B-02-C	44	1.00	2.00	2.00	100.0
K3927-03	03-A-03-B-03-C	45	1.00	2.00	2.00	100.0
K3927-04	04-A-04-B-04-C	46	1.00	2.00	2.00	100.0
K3928-01	ROLL-OFF-COMP	47	1.13	9.77	8.81	88.9
K3928-02	ROLL-OFF-COMP-A	48	1.13	9.77	8.81	88.9
K3928-03	ROLL-OFF-COMP-B	49	1.13	9.77	8.81	88.9
K3928-04	ROLL-OFF-COMP	50	1.13	9.77	8.81	88.9
K3929-03	30319	51	1.00	2.00	2.00	100.0
K3929-05	30320	52	1.00	2.00	2.00	100.0
K3932-01	390718084	53	1.18	9.98	3.04	21.1
K3933-02	20190719-WWT-1	54	1.14	9.76	5.18	46.9
K3933-03	20190719-WWT-2	55	1.13	9.75	6.1	57.7
K3933-04	20190719-WWT-3	56	1.14	9.67	5.85	55.2
K3933-05	20190719-PLATING-31	57	1.14	9.8	8.5	85.0
K3933-06	20190719-PLATING-32	58	1.14	9.72	7.39	72.8
K3933-07	20190719-PLATING-33	59	1.14	9.55	8.63	89.1
K3933-08	20190719-PLATING-COMP	60	1.13	9.98	8.49	83.2
K3933-09	20190719-WWT-COMP	61	1.13	9.78	5.59	51.6

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$