



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104103

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/19/2019	18:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/19/2019	18:01
3	CAL3	1	Water	NA	NA	20.2	10.03	07/19/2019	18:02
4	ICV	1	Water	NA	NA	20.2	6.99	07/19/2019	18:04
5	CCV1	1	Water	NA	NA	20.2	2.02	07/19/2019	18:05
6	K3617-02	1	Solid	20.02	20	23.4	7.86	07/19/2019	18:07
7	K3617-04	1	Solid	20.03	20	23.3	7.84	07/19/2019	18:10
8	K3621-06	1	Solid	20.02	20	24.5	8.92	07/19/2019	18:11
9	K3622-10	1	Solid	20.04	20	24.6	8.73	07/19/2019	18:12
10	K3622-12	1	Solid	20.02	20	24.4	8.64	07/19/2019	18:15
11	K3891-07	1	Solid	20.02	20	24.1	5.11	07/19/2019	18:17
12	K3896-06	1	Solid	20.01	20	24.4	4.31	07/19/2019	18:20
13	K3896-07	1	Solid	20.02	20	24.7	4.50	07/19/2019	18:21
14	K3896-08	1	Solid	20.04	20	24.7	4.47	07/19/2019	18:22
15	K3896-09	1	Solid	20.03	20	24.2	4.55	07/19/2019	18:24
16	CCV2	1	Water	NA	NA	20.1	12.02	07/19/2019	18:25
17	K3928-04	1	Solid	20.04	20	23.7	9.32	07/19/2019	18:26
18	K3928-04DUP	1	Solid	20.03	20	23.8	9.33	07/19/2019	18:27
19	CCV3	1	Water	NA	NA	20.2	2.02	07/19/2019	18:30

LB 104103

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K3617 WorkList ID : 128808 Department : Wet-Chemistry Date : 07-19-2019 15:56:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/19/2019	Solid	K3617-02 -1	Corrosivity	Cool 4 deg C	PSEG05	M51	NB-07-071919-A	07/19/2019	9045D
07/19/2019	Solid	K3617-04 -1	Corrosivity	Cool 4 deg C	PSEG05	M51	NB-07-071919-B	07/19/2019	9045D
07/19/2019	Solid	K3621-06 -1	Corrosivity	Cool 4 deg C	PSEG05		OK-02-071919	07/19/2019	9045D
07/19/2019	Solid	K3622-10 -1	Corrosivity	Cool 4 deg C	PSEG05		OR-02-071919-A	07/19/2019	9045D
07/19/2019	Solid	K3622-12 -1	Corrosivity	Cool 4 deg C	PSEG05		OR-02-071919-B	07/19/2019	9045D
07/16/2019	Solid	K3891-07 -1	Corrosivity	Cool 4 deg C	ROYF02	M63	P001-SW03-01	07/16/2019	9045D
07/17/2019	Solid	K3896-06 -1	Corrosivity	Cool 4 deg C	SKAN04	M62	TRAN-134	07/17/2019	9045D
07/17/2019	Solid	K3896-07 -1	Corrosivity	Cool 4 deg C	SKAN04	M62	TRAN-135	07/17/2019	9045D
07/17/2019	Solid	K3896-08 -1	Corrosivity	Cool 4 deg C	SKAN04	M62	TRAN-105	07/17/2019	9045D
07/17/2019	Solid	K3896-09 -1	Corrosivity	Cool 4 deg C	SKAN04	M62	TRAN-104	07/17/2019	9045D
07/19/2019	Solid	K3928-04 -1	Corrosivity	Cool 4 deg C	PSEG01	M63	ROLL-OFF-COMP	07/19/2019	9045D

Date/Time 07-19-19 16:15

Received by: JH

Relinquished by: RC

Date/Time 07-19-19 16:30

Received by: JH

Relinquished by: JH