



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104104

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/19/2019	18:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/19/2019	18:01
3	CAL3	1	Water	NA	NA	20.2	10.03	07/19/2019	18:02
4	ICV	1	Water	NA	NA	20.2	6.99	07/19/2019	18:04
5	CCV1	1	Water	NA	NA	20.2	2.02	07/19/2019	18:05
6	K3931-01	1	Water	NA	NA	24.1	5.87	07/19/2019	18:32
7	K3931-01DUP	1	Water	NA	NA	24.2	5.88	07/19/2019	18:33
8	CCV2	1	Water	NA	NA	20.1	12.02	07/19/2019	18:35

18104104

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3931

WorkList ID : 128809

Department : Wet-Chemistry

Date : 07-19-2019 15:56:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/15/2019	Water	K3931-01- 1	pH	Cool 4 deg C	STER10	M61	CELL-A-POOL-WATER(DE-IC	07/15/2019	SM 4500-H B

Date/Time 07-19-19 16:00
Received by: SA
Relinquished by: SA

Date/Time 07-19-19 17:00
Received by: SA
Relinquished by: SA