



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104105

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/19/2019	18:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/19/2019	18:01
3	CAL3	1	Water	NA	NA	20.2	10.03	07/19/2019	18:02
4	ICV	1	Water	NA	NA	20.2	6.99	07/19/2019	18:04
5	CCV1	1	Water	NA	NA	20.1	2.02	07/19/2019	18:35
6	K3926-01	1	Water	NA	NA	27.0	8.10	07/19/2019	18:36
7	K3926-01DUP	1	Water	NA	NA	27.1	8.11	07/19/2019	18:37
8	K3928-05	1	Water	NA	NA	22.8	7.07	07/19/2019	18:39
9	K3928-06	1	Water	NA	NA	23.2	7.10	07/19/2019	18:40
10	K3928-07	1	Water	NA	NA	23.3	7.04	07/19/2019	18:41
11	K3928-08	1	Water	NA	NA	22.9	7.12	07/19/2019	18:43
12	K3929-10	1	Water	NA	NA	23.5	8.76	07/19/2019	18:45
13	K3929-11	1	Water	NA	NA	22.9	5.70	07/19/2019	18:46
14	K3929-12	1	Water	NA	NA	23.3	9.54	07/19/2019	18:47
15	K3929-13	1	Water	NA	NA	22.6	9.64	07/19/2019	18:49
16	CCV2	1	Water	NA	NA	20.2	12.02	07/19/2019	18:50
17	K3929-14	1	Water	NA	NA	22.7	8.80	07/19/2019	18:51
18	CCV3	1	Water	NA	NA	20.1	2.02	07/19/2019	18:52

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3926 WorkList ID : 128806 Department : Wet-Chemistry Date : 07-19-2019 15:54:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/19/2019	Water	K3926-01	pH	Cool 4 deg C	PSEG01	M63	POLY-TANK-GROUNDWATER	07/19/2019	9040C
07/19/2019	Water	K3928-05	pH	Cool 4 deg C	PSEG01	M63	30380	07/19/2019	9040C
07/19/2019	Water	K3928-06	pH	Cool 4 deg C	PSEG01	M63	30381	07/19/2019	9040C
07/19/2019	Water	K3928-07	pH	Cool 4 deg C	PSEG01	M63	30382	07/19/2019	9040C
07/19/2019	Water	K3928-08	pH	Cool 4 deg C	PSEG01	M63	TOTE-COMP	07/19/2019	9040C
07/19/2019	Water	K3929-10	pH	Cool 4 deg C	PSEG01	M63	30302	07/19/2019	9040C
07/19/2019	Water	K3929-11	pH	Cool 4 deg C	PSEG01	M63	30283	07/19/2019	9040C
07/19/2019	Water	K3929-12	pH	Cool 4 deg C	PSEG01	M63	30377	07/19/2019	9040C
07/19/2019	Water	K3929-13	pH	Cool 4 deg C	PSEG01	M63	30378	07/19/2019	9040C
07/19/2019	Water	K3929-14	pH	Cool 4 deg C	PSEG01	M63	30379	07/19/2019	9040C

Date/Time 07-19-19 16:00
 Received by: SO
 Relinquished by: SO

Date/Time 07-19-19 17:35
 Received by: SO
 Relinquished by: SO