



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104106

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/19/2019	18:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/19/2019	18:01
3	CAL3	1	Water	NA	NA	20.2	10.03	07/19/2019	18:02
4	ICV	1	Water	NA	NA	20.2	6.99	07/19/2019	18:04
5	CCV1	1	Water	NA	NA	20.1	2.02	07/19/2019	18:35
6	K3891-01	1	Water	NA	NA	23.9	5.90	07/19/2019	18:36
7	K3891-02	1	Water	NA	NA	23.7	5.23	07/19/2019	18:27
8	K3891-04	1	Water	NA	NA	23.6	5.08	07/19/2019	18:29
9	K3891-05	1	Water	NA	NA	24.1	5.28	07/19/2019	18:30
10	K3891-06	1	Water	NA	NA	24.1	5.19	07/19/2019	18:31
11	K3926-02	1	Water	NA	NA	27.0	8.10	07/19/2019	18:33
12	K3926-02DUP	1	Water	NA	NA	27.1	8.10	07/19/2019	18:34
13	K3928-09	1	Water	NA	NA	22.9	7.12	07/19/2019	18:35
14	CCV2	1	Water	NA	NA	20.1	12.02	07/19/2019	18:52

2B 104106

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K3926 WorkList ID : 128807 Department : Wet-Chemistry Date : 07-19-2019 15:55:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/16/2019	Water	K3891-01	Corrosivity	Cool 4 deg C	ROYF02	M62	P001-SW01-01	07/16/2019	9040C
07/16/2019	Water	K3891-02	Corrosivity	Cool 4 deg C	ROYF02	M62	P001-SW02-01	07/16/2019	9040C
07/16/2019	Water	K3891-04	Corrosivity	Cool 4 deg C	ROYF02	M62	P001-SW04-01	07/16/2019	9040C
07/16/2019	Water	K3891-05	Corrosivity	Cool 4 deg C	ROYF02	M62	P001-SW05-01	07/16/2019	9040C
07/16/2019	Water	K3891-06	Corrosivity	Cool 4 deg C	ROYF02	M62	P001-SW06-01	07/16/2019	9040C
07/19/2019	Water	K3926-02	Corrosivity	Cool 4 deg C	PSEG01	M63	POLY-TANK-GROUNDWATER	07/19/2019	9040C
07/19/2019	Water	K3928-09	Corrosivity	Cool 4 deg C	PSEG01	M63	TOTE-COMP	07/19/2019	9040C

Date/Time 07-19-19 16:00
 Received by: Jn L.C.
 Relinquished by:

Date/Time 07-19-19 18:30
 Received by: SD
 Relinquished by: 32