



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104107

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/19/2019	18:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/19/2019	18:01
3	CAL3	1	Water	NA	NA	20.2	10.03	07/19/2019	18:02
4	ICV	1	Water	NA	NA	20.2	6.99	07/19/2019	18:04
5	CCV1	1	Water	NA	NA	20.2	2.01	07/19/2019	19:00
6	K3929-03	1	Solid	20.02	20	24.1	10.68	07/19/2019	19:02
7	K3929-05	1	Solid	20.03	20	27.4	7.30	07/19/2019	19:03
8	K3929-06	1	Solid	20.04	20	24.5	5.04	07/19/2019	19:05
9	K3929-06DUP	1	Solid	20.03	20	24.6	5.05	07/19/2019	19:06
10	CCV2	1	Water	NA	NA	20.2	12.02	07/19/2019	19:10

28704107

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3926 S WorkList ID : 128811 Department : Wet-Chemistry Date : 07-19-2019 16:15:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/19/2019	Solid	K3929-03	pH	Cool 4 deg C	PSEG01	M63	30319	07/19/2019	9045D
07/19/2019	Solid	K3929-05	pH	Cool 4 deg C	PSEG01	M63	30320	07/19/2019	9045D
07/19/2019	Solid	K3929-06	pH	Cool 4 deg C	PSEG01	M63	30321	07/19/2019	9045D

Date/Time 07-19-19 16:20
Received by: J
Relinquished by: J

Date/Time 07-19-19 17:40
Received by: J
Relinquished by: J