



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104139

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/22/2019	11:02
2	K3896-06		1	8	100.00	>212.0	>212.0	07/22/2019	11:32
3	K3896-06DUP		1	8	100.00	>212.0	>212.0	07/22/2019	11:50
4	K3896-07		1	9	100.00	>212.0	>212.0	07/22/2019	12:15
5	K3896-08		1	8	100.00	>212.0	>212.0	07/22/2019	12:40
6	K3896-09		1	7	100.00	>212.0	>212.0	07/22/2019	13:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 104139

WorkList Name : FP72219 WorkList ID : 128866 Department : Wet-Chemistry Date : 07-22-2019 10:36:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/19/2019	Solid	K3896-06 ✓	Flash Point	Cool 4 deg C	SKAN04	M62	TRAN-134	07/17/2019	1010A
07/19/2019	Solid	K3896-07 ✓	Flash Point	Cool 4 deg C	SKAN04	M62	TRAN-135	07/17/2019	1010A
07/19/2019	Solid	K3896-08 ✓	Flash Point	Cool 4 deg C	SKAN04	M62	TRAN-105	07/17/2019	1010A
07/19/2019	Solid	K3896-09 ✓	Flash Point	Cool 4 deg C	SKAN04	M62	TRAN-104	07/17/2019	1010A

Date/Time 07/22/19 10:40
 Received by: KS
 Relinquished by: CP

Date/Time 07/22/19 13:10
 Received by: CP
 Relinquished by: KS