



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104144

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/22/2019	14:20
2	CAL2	1	Water	NA	NA	20.1	6.99	07/22/2019	14:21
3	CAL3	1	Water	NA	NA	20.1	10.03	07/22/2019	14:22
4	ICV	1	Water	NA	NA	20.1	7.01	07/22/2019	14:24
5	CCV1	1	Water	NA	NA	20.1	2.02	07/22/2019	14:25
6	K3929-16	1	Water	NA	NA	27.4	7.30	07/22/2019	14:26
7	K3934-13	1	Water	NA	NA	18.6	7.59	07/22/2019	14:27
8	K3934-16	1	Water	NA	NA	18.7	7.60	07/22/2019	14:29
9	K3934-16DUP	1	Water	NA	NA	18.7	7.60	07/22/2019	14:30
10	CCV2	1	Water	NA	NA	20.2	12.02	07/22/2019	14:31

2B 10h/4h

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K3934 WorkList ID : 128888 Department : Wet-Chemistry Date : 07-22-2019 14:12:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/19/2019	Water	K3929-16	pH	Cool 4 deg C	PSEG01	M63	30320	07/19/2019	9040C
07/19/2019	Water	K3934-13	pH	Cool 4 deg C	PSEG01	M63	TP-1	07/19/2019	9040C
07/19/2019	Water	K3934-16	pH	Cool 4 deg C	PSEG01	M63	TP-3	07/19/2019	9040C

Date/Time 07-22-19 14:15
Received by: ga
Relinquished by: ga

Date/Time 07-22-19 15:00
Received by: ga
Relinquished by: ga