


LB104157

7470A

INSTRUMENT ID : C U 1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.036	0.036	138	0.000	138				
0.2	0.200	0.200	0.000	802	0.0 %	802				
2.5	2.500	2.362	-0.138	9523	0.0 %	9523				
5.0	5.000	5.069	0.069	20450	0.0 %	20450				
7.5	7.500	7.660	0.160	30902	0.0 %	30902				
10.0	10.000	9.873	-0.127	39834	0.0 %	39834				

LB104157 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Jul 2019 10:49:20

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	138	-	0.0007470A	PPB		23 Jul 2019 10:53:06	S	Std
0.2 - 1	50-2	802	-	0.2007470A	PPB		23 Jul 2019 10:55:15	S	Std
2.5 - 1	52-5	9523	-	2.5007470A	PPB		23 Jul 2019 10:57:24	S	Std
5.0 - 1	55-0	20450	-	5.0007470A	PPB		23 Jul 2019 10:59:32	S	Std
7.5 - 1	57-5	30902	-	7.5007470A	PPB		23 Jul 2019 11:01:40	S	Std
10.0 - 1	510-0	39834	-	10.0007470A	PPB		23 Jul 2019 11:06:26	S	Std
ICV42 - 1	ICV42	15599	3.8673	-7470A	PPB		23 Jul 2019 11:09:32	U	SMPL
ICB42 - 1	ICB42	279	0.0706	-7470A	PPB		23 Jul 2019 11:11:38	U	SMPL
CCV67 - 1	CCV67	20807	5.1580	-7470A	PPB		23 Jul 2019 11:13:45	U	SMPL
CCB67 - 1	CCB67	83	0.0220	-7470A	PPB		23 Jul 2019 11:15:52	U	SMPL
CRA - 1	CRA	796	0.1987	-7470A	PPB		23 Jul 2019 11:17:58	U	SMPL
HighStd - 1	HighStd	40410	10.0161	-7470A	PPB		23 Jul 2019 11:20:04	U	SMPL
ChkStd - 1	ChkStd	29774	7.3882	-7470A	PPB		23 Jul 2019 11:22:11	U	SMPL
PB121647BL - 1	PBW	0	0.0015	-7470A	PPB		23 Jul 2019 11:24:19	U	SMPL
ChkStd - 1	ChkStd	28720	7.1190	-7470A	PPB		23 Jul 2019 11:25:46	U	SMPL
PB121647BL - 1	PBW	111	0.0290	-7470A	PPB		23 Jul 2019 11:27:53	U	SMPL
PB121647BS - 1	LCSW	15564	3.8586	-7470A	PPB		23 Jul 2019 11:30:00	U	SMPL
K3617-02 - 1	NB-07-071919-A	153	0.0394	-7470A	PPB		23 Jul 2019 11:32:09	U	SMPL
K3617-02DUP - 1	NB-07-071919-ADUP	183	0.0468	-7470A	PPB		23 Jul 2019 11:34:17	U	SMPL
K3617-02MS - 1	NB-07-071919-AMS	16440	4.0757	-7470A	PPB		23 Jul 2019 11:36:25	U	SMPL
K3617-02MSD - 1	NB-07-071919-AMSD	16719	4.1449	-7470A	PPB		23 Jul 2019 11:38:33	U	SMPL
K3617-04 - 1	NB-07-071919-B	394	0.0991	-7470A	PPB		23 Jul 2019 11:40:43	U	SMPL
CCV68 - 1	CCV68	20532	5.0898	-7470A	PPB		23 Jul 2019 11:42:50	U	SMPL
CCB68 - 1	CCB68	120	0.0312	-7470A	PPB		23 Jul 2019 11:44:56	U	SMPL
K3621-06 - 1	OK-02-071919	199	0.0508	-7470A	PPB		23 Jul 2019 11:47:03	U	SMPL
K3622-10 - 1	OR-02-071919-A	190	0.0485	-7470A	PPB		23 Jul 2019 11:49:12	U	SMPL
K3622-12 - 1	OR-02-071919-B	177	0.0453	-7470A	PPB		23 Jul 2019 11:51:22	U	SMPL
K3928-04 - 1	ROLL-OFF-COMP	169	0.0433	-7470A	PPB		23 Jul 2019 11:53:29	U	SMPL
K3929-06 - 1	30321	194	0.0495	-7470A	PPB		23 Jul 2019 11:55:36	U	SMPL
PB121648BL - 1	PBW	194	0.0495	-7470A	PPB		23 Jul 2019 11:57:44	U	SMPL
PB121648BS - 1	LCSW	15862	3.9325	-7470A	PPB		23 Jul 2019 11:59:52	U	SMPL
K3901-02 - 1	CONTAINMENT-WATER	183	0.0468	-7470A	PPB		23 Jul 2019 12:02:00	U	SMPL
K3926-02 - 1	POLY-TANK-GROUNDWATER	202	0.0515	-7470A	PPB		23 Jul 2019 12:04:09	U	SMPL
K3926-02DUP - 1	POLY-TANK-GROUNDWATERDUP	190	0.0485	-7470A	PPB		23 Jul 2019 12:06:16	U	SMPL
CCV69 - 1	CCV69	20583	5.1025	-7470A	PPB		23 Jul 2019 12:08:23	U	SMPL
CCB69 - 1	CCB69	123	0.0319	-7470A	PPB		23 Jul 2019 12:10:29	U	SMPL
K3926-02MS - 1	POLY-TANK-GROUNDWATERMS	16132	3.9994	-7470A	PPB		23 Jul 2019 12:12:36	U	SMPL
K3926-02MSD - 1	POLY-TANK-GROUNDWATERMSD	16605	4.1166	-7470A	PPB		23 Jul 2019 12:14:43	U	SMPL
K3928-09 - 1	TOTE-COMP	173	0.0443	-7470A	PPB		23 Jul 2019 12:16:50	U	SMPL
PB121649BL - 1	PBW	176	0.0451	-7470A	PPB		23 Jul 2019 12:18:56	U	SMPL
PB121649BS - 1	LCSW	15896	3.9409	-7470A	PPB		23 Jul 2019 12:21:04	U	SMPL
K2646-15 - 1	COMP-3	3144	0.7806	-7470A	PPB		23 Jul 2019 12:23:11	U	SMPL
K3891-01 - 1	P001-SW01-01	355	0.0894	-7470A	PPB		23 Jul 2019 12:25:20	U	SMPL
K3891-02 - 1	P001-SW02-01	157	0.0404	-7470A	PPB		23 Jul 2019 12:27:28	U	SMPL
K3891-04 - 1	P001-SW04-01	2110	0.5244	-7470A	PPB		23 Jul 2019 12:29:36	U	SMPL
K3891-05 - 1	P001-SW05-01	-65	-0.0147	-7470A	PPB		23 Jul 2019 12:31:43	U	SMPL
CCV70 - 1	CCV70	19065	4.7263	-7470A	PPB		23 Jul 2019 12:58:31	U	SMPL
CCB70 - 1	CCB70	117	0.0305	-7470A	PPB		23 Jul 2019 13:00:38	U	SMPL
K3891-06 - 1	P001-SW06-01	64	0.0173	-7470A	PPB		23 Jul 2019 13:02:46	U	SMPL
K3926-01 - 1	POLY-TANK-GROUNDWATER	5267	1.3068	-7470A	PPB		23 Jul 2019 13:04:52	U	SMPL
K3928-08 - 1	TOTE-COMP	250	0.0634	-7470A	PPB		23 Jul 2019 13:06:59	U	SMPL
K3928-08DUP - 1	TOTE-COMPDUP	269	0.0681	-7470A	PPB		23 Jul 2019 13:09:05	U	SMPL
K3928-08MS - 1	TOTE-COMPMS	14684	3.6405	-7470A	PPB		23 Jul 2019 13:11:13	U	SMPL
K3928-08MSD - 1	TOTE-COMPMSD	15818	3.9216	-7470A	PPB		23 Jul 2019 13:13:20	U	SMPL
K3934-13 - 1	TP-1	2040	0.5070	-7470A	PPB		23 Jul 2019 13:15:28	U	SMPL
K3934-16 - 1	TP-3	8733	2.1657	-7470A	PPB		23 Jul 2019 13:17:39	U	SMPL
PB121647TB - 1	PB121555TB	121	0.0314	-7470A	PPB		23 Jul 2019 13:19:49	U	SMPL
PB121648TB - 1	PB121557TB	153	0.0394	-7470A	PPB		23 Jul 2019 13:22:00	U	SMPL
CCV71 - 1	CCV71	18542	4.5966	-7470A	PPB		23 Jul 2019 13:24:14	U	SMPL
CCB71 - 1	CCB71	135	0.0349	-7470A	PPB		23 Jul 2019 13:26:20	U	SMPL
K3617-02LX5 - 1		121	0.0314	-7470A	PPB		23 Jul 2019 13:28:27	U	SMPL
K3617-02A - 1		16454	4.0792	-7470A	PPB		23 Jul 2019 13:30:37	U	SMPL
K3926-02LX5 - 1		94	0.0248	-7470A	PPB		23 Jul 2019 13:32:46	U	SMPL
K3926-02A - 1		15894	3.9404	-7470A	PPB		23 Jul 2019 13:34:56	U	SMPL
K3928-08LX5 - 1		83	0.0220	-7470A	PPB		23 Jul 2019 13:37:14	U	SMPL
K3928-08A - 1		15675	3.8861	-7470A	PPB		23 Jul 2019 13:39:26	U	SMPL
CCV72 - 1	CCV72	19795	4.9072	-7470A	PPB		23 Jul 2019 13:41:36	U	SMPL
CCB72 - 1	CCB72	19	0.0062	-7470A	PPB		23 Jul 2019 13:43:45	U	SMPL

NS 07-23-19