


LB104163

7471B

INSTRUMENT ID: CU1

Linear



A= 0.0000e+000
B= 1.8050e-004
C= 3.3569e-002
Rho= 0.9999070
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.051	0.051	97	0.000	97				
0.2	0.200	0.223	0.023	1048	0.0 %	1048				
2.5	2.500	2.465	-0.035	13471	0.0 %	13471				
5.0	5.000	4.916	-0.084	27048	0.0 %	27048				
7.5	7.500	7.486	-0.014	41284	0.0 %	41284				
10.0	10.000	10.060	0.060	55545	0.0 %	55545				

LB104163 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 23 Jul 2019 14:38:27					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	97	-	0.0000	7471B	PPB	23 Jul 2019 14:39:59	S	Std
0.2 - 1	50.2	1048	-	0.2000	7471B	PPB	23 Jul 2019 14:42:17	S	Std
2.5 - 1	52.5	13471	-	2.5000	7471B	PPB	23 Jul 2019 14:44:35	S	Std
5.0 - 1	55.0	27048	-	5.0000	7471B	PPB	23 Jul 2019 14:46:53	S	Std
7.5 - 1	57.5	41284	-	7.5000	7471B	PPB	23 Jul 2019 14:49:11	S	Std
10.0 - 1	510.0	55545	-	10.0000	7471B	PPB	23 Jul 2019 15:03:08	S	Std
ICV43 - 1	ICV43	21953	3.9962	-	7471B	PPB	23 Jul 2019 15:07:02	U	SMPL
ICB43 - 1	ICB43	53	0.0431	-	7471B	PPB	23 Jul 2019 15:09:17	U	SMPL
CCV73 - 1	CCV73	26857	4.8814	-	7471B	PPB	23 Jul 2019 15:11:33	U	SMPL
CCB73 - 1	CCB73	36	0.0401	-	7471B	PPB	23 Jul 2019 15:13:49	U	SMPL
CRA - 1	CRA	1013	0.2164	-	7471B	PPB	23 Jul 2019 15:16:06	U	SMPL
HighStd - 1	HighStd	55092	9.9779	-	7471B	PPB	23 Jul 2019 15:18:22	U	SMPL
ChkStd - 1	ChkStd	38044	6.9007	-	7471B	PPB	23 Jul 2019 15:20:38	U	SMPL
PB121651BL - 1	PBS	33	0.0395	-	7471B	PPB	23 Jul 2019 15:23:03	U	SMPL
PB121651BS - 1	LCSS	19029	3.4684	-	7471B	PPB	23 Jul 2019 15:25:21	U	SMPL
K3617-01 - 1	NB-07-071919-A	1480	0.3007	-	7471B	PPB	23 Jul 2019 15:27:39	U	SMPL
K3617-01DUP - 1	NB-07-071919-ADUP	1497	0.3038	-	7471B	PPB	23 Jul 2019 15:29:58	U	SMPL
K3617-01MS - 1	NB-07-071919-AMS	20446	3.7242	-	7471B	PPB	23 Jul 2019 15:32:15	U	SMPL
K3617-01MSD - 1	NB-07-071919-AMSD	20668	3.7642	-	7471B	PPB	23 Jul 2019 15:34:33	U	SMPL
K3617-03 - 1	NB-07-071919-B	1376	0.2819	-	7471B	PPB	23 Jul 2019 15:36:50	U	SMPL
CCV74 - 1	CCV74	26558	4.8274	-	7471B	PPB	23 Jul 2019 15:39:08	U	SMPL
CCB74 - 1	CCB74	31	0.0392	-	7471B	PPB	23 Jul 2019 15:41:24	U	SMPL
K3618-07 - 1	HD-02-072219	1556	0.3144	-	7471B	PPB	23 Jul 2019 15:43:41	U	SMPL
K3621-05 - 1	OK-02-071919	1052	0.2235	-	7471B	PPB	23 Jul 2019 15:45:57	U	SMPL
K3622-09 - 1	OR-02-071919-A	2029	0.3998	-	7471B	PPB	23 Jul 2019 15:48:14	U	SMPL
K3622-11 - 1	OR-02-071919-B	1834	0.3646	-	7471B	PPB	23 Jul 2019 15:50:30	U	SMPL
K3891-07 - 1	P001-SW03-01	267	0.0818	-	7471B	PPB	23 Jul 2019 15:52:47	U	SMPL
K3916-01 - 1	GIS-BUILDING-1	538	0.1307	-	7471B	PPB	23 Jul 2019 15:55:04	U	SMPL
K3916-02 - 1	GIS-BUILDING-2	545	0.1319	-	7471B	PPB	23 Jul 2019 15:57:21	U	SMPL
K3916-03 - 1	GIS-BUILDING-3	269	0.0821	-	7471B	PPB	23 Jul 2019 15:59:39	U	SMPL
K3928-01 - 1	ROLL-OFF-COMP	2628	0.5079	-	7471B	PPB	23 Jul 2019 16:01:57	U	SMPL
K3933-08 - 1	20190719-PLATING-COMP	16570	3.0245	-	7471B	PPB	23 Jul 2019 16:04:13	U	SMPL
CCV75 - 1	CCV75	26543	4.8247	-	7471B	PPB	23 Jul 2019 16:06:29	U	SMPL
CCB75 - 1	CCB75	67	0.0457	-	7471B	PPB	23 Jul 2019 16:08:47	U	SMPL
K3933-09 - 1	20190719-WWT-COMP	969	0.2085	-	7471B	PPB	23 Jul 2019 16:11:05	U	SMPL
K3934-03 - 1	TP-2S	14334	2.6209	-	7471B	PPB	23 Jul 2019 16:13:21	U	SMPL
K3934-07 - 1	TP-3S	78553	14.2127	-	7471B	PPB	23 Jul 2019 16:15:38	U	SMPL
K3934-11 - 1	TP-1S	7192	1.3318	-	7471B	PPB	23 Jul 2019 16:17:54	U	SMPL
K3936-01 - 1	SOIL-COMP	1953	0.3861	-	7471B	PPB	23 Jul 2019 16:20:14	U	SMPL
K3943-01 - 1	PA-71619	5653	1.0540	-	7471B	PPB	23 Jul 2019 16:22:31	U	SMPL
K3617-01LX5 - 1		355	0.0976	-	7471B	PPB	23 Jul 2019 16:24:48	U	SMPL
K3617-01A - 1		20613	3.7543	-	7471B	PPB	23 Jul 2019 16:27:06	U	SMPL
K3934-07DLX5 - 1		16270	2.9704	-	7471B	PPB	23 Jul 2019 16:29:24	U	SMPL
CCV76 - 1	CCV76	26923	4.8933	-	7471B	PPB	23 Jul 2019 16:35:11	U	SMPL
CCB76 - 1	CCB76	-19	0.0301	-	7471B	PPB	23 Jul 2019 16:37:27	U	SMPL