



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104173

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/23/2019	16:25
2	CAL2	1	Water	NA	NA	20.1	6.99	07/23/2019	16:26
3	CAL3	1	Water	NA	NA	20.1	10.03	07/23/2019	16:27
4	ICV	1	Water	NA	NA	20.1	7.01	07/23/2019	16:30
5	CCV1	1	Water	NA	NA	20.2	2.02	07/23/2019	16:32
6	K3969-03	1	Water	NA	NA	13.2	7.08	07/23/2019	16:35
7	K3969-03DUP	1	Water	NA	NA	13.3	7.09	07/23/2019	16:36
8	CCV2	1	Water	NA	NA	20.1	12.02	07/23/2019	16:37

28104173

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K3969 WorkList ID : 128939 Department : Wet-Chemistry Date : 07-23-2019 16:20:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/23/2019	Water	K3969-03 -P	Corrosivity	Cool 4 deg C	DVIR01	M61	TANK-1	07/23/2019	9040C

Date/Time 07-23-19 16:22
 Received by: JH
 Relinquished by: CD

Date/Time 07-23-19 17:00
 Received by: CD
 Relinquished by: J3