



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB104176

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/22/2019	11:20
2	K3891-01		1	8	100.00	>212.0	>212.0	07/22/2019	11:40
3	K3891-01DUP		1	8	100.00	>212.0	>212.0	07/22/2019	12:05
4	K3891-02		1	9	100.00	>212.0	>212.0	07/22/2019	12:35
5	K3891-04		1	7	100.00	>212.0	>212.0	07/22/2019	13:05
6	K3891-05		1	8	100.00	>212.0	>212.0	07/22/2019	13:40
7	K3891-06		1	7	100.00	>212.0	>212.0	07/22/2019	14:10
8	K3926-02		1	9	100.00	>212.0	>212.0	07/22/2019	14:45
9	K3928-09		1	9	100.00	>212.0	>212.0	07/22/2019	15:10
10	K3929-01		1	7	100.00	>212.0	>212.0	07/22/2019	15:40
11	K3929-08		1	7	100.00	>212.0	>212.0	07/22/2019	16:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 104176

WorkList Name : FP72219-1 WorkList ID : 12867 Department : Wet-Chemistry Date : 07-22-2019 10:38:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/30/2019	Water	K3891-01	B Flash Point	Cool 4 deg C	ROYF02	M62	P001-SW01-01	07/16/2019	1010A
07/30/2019	Water	K3891-02	B Flash Point	Cool 4 deg C	ROYF02	M62	P001-SW02-01	07/16/2019	1010A
07/30/2019	Water	K3891-04	B Flash Point	Cool 4 deg C	ROYF02	M62	P001-SW04-01	07/16/2019	1010A
07/30/2019	Water	K3891-05	B Flash Point	Cool 4 deg C	ROYF02	M62	P001-SW05-01	07/16/2019	1010A
07/30/2019	Water	K3891-06	B Flash Point	Cool 4 deg C	ROYF02	M62	P001-SW06-01	07/16/2019	1010A
07/24/2019	Water	K3926-02	C Flash Point	Cool 4 deg C	PSEG01	M63	POLY-TANK-GROUNDWATER	07/19/2019	1010A
07/24/2019	Water	K3928-09	C Flash Point	Cool 4 deg C	PSEG01	M63	TOTE-COMP	07/19/2019	1010A
07/26/2019	Water	K3929-01	B Flash Point	Cool 4 deg C	PSEG01	M63	30302-30283	07/19/2019	1010A
07/26/2019	Water	K3929-08	A Flash Point	Cool 4 deg C	PSEG01	M63	30377-78-79	07/19/2019	1010A

Date/Time 07/22/19 10:45
 Received by: KS
 Relinquished by: CP

Date/Time 07/22/19 16:15
 Received by: CP
 Relinquished by: KS