

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 07/25/2019

OVEN TEMP1: 98 °C TEMP: 104 °C

DATE & TIME IN: 07/24/2019 15:35 OUT: 07/25/2019 07:35

OVEN TEMP2: 104 °C TEMP: 104 °C

DATE & TIME IN: 07/25/2019 09:40 OUT: 07/25/2019 12:50

Run Number: LB104190

BalanceID: WC SC-3

OvenID: Wet Oven-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB104190BL	LB104190BL	95.2037	95.2037	100	95.2037	95.2037	95.2037	0	0
2	K3992-01	55-ST-19	90.0031	90.0031	100	90.1364	90.1364	90.1364	0.1333	1333
3	K3992-01DUP	55-ST-19DUP	90.0673	90.0673	100	90.2009	90.2009	90.2009	0.1336	1336

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

283104190

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS K3992 WorkList ID : 128985 Department : Wet-Chemistry Date : 07-24-2019 15:19:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/31/2019	Water	K3992-01	TS	Cool 4 deg C	DIST01	N11	55-ST-19	07/24/2019	SM2540 B

Date/Time 07-24-14 15:20
Received by: Jn
Relinquished by: CB

Date/Time 07-24-14 16:30
Received by: CB
Relinquished by: Jn