



Analytical Summary Report

Analysis Method: SM2130 B  
Parameter: Turbidity  
Run Number: LB104191

ANALYST: AMANDEEP  
SUPERVISOR REVIEW BY: Apatel

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP76712
Turbidity Calibration std, 0NTU	WP76711
Turbidity Calibration std, 20NTU	WP76707
Turbidity Calibration std, 80NTU	WP76705
Turbidity Calibration std, 40NTU	WP76706
Turbidity Calibration std, 200NTU	WP76704
Turbidity Calibration std, 1NTU	WP76710
Turbidity Calibration std, 10NTU	WP76708
Turbidity Calibration std, 5NTU	WP76709
10 NTU Standard 500 ml	W2297

Intercept: 0.5434                      Slope: 0.9053                      Regression: 0.9995

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.310	07/24/2019	10:31
2	CAL2	1	1	0.996	07/24/2019	10:38
3	CAL3	5	1	4.820	07/24/2019	10:45
4	CAL4	10	1	9.600	07/24/2019	10:52
5	CAL5	20	1	18.700	07/24/2019	10:58
6	CAL6	40	1	38.500	07/24/2019	11:05
7	CAL7	80	1	72.100	07/24/2019	12:10



Analytical Summary Report

Analysis Method: SM2130 B  
Parameter: Turbidity  
Run Number: LB104191

ANALYST: AMANDEEP  
SUPERVISOR REVIEW BY: Apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.450	9.838	07/24/2019	11:13
2	ICB		1	0.298	-0.271	07/24/2019	11:17
3	CCV1	10	1	9.520	9.916	07/24/2019	11:25
4	CCB1		1	0.296	-0.273	07/24/2019	11:30
5	LB104191BL		1	0.301	-0.268	07/24/2019	11:36
6	K3971-01		1	3.850	3.652	07/24/2019	11:41
7	K3972-01		1	56.100	61.368	07/24/2019	11:47
8	K3972-01DUP		1	56.000	61.258	07/24/2019	11:52
9	CCV2	10	1	9.560	9.960	07/24/2019	11:58
10	CCB2		1	0.285	-0.285	07/24/2019	12:03

WORKLIST(Hardcopy Internal Chain)

LB104191

WorkList Name : TUR072419      WorkList ID : 128972      Department : Wet-Chemistry      Date : 07-24-2019 09:10:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/25/2019	Water	K3971-01	Turbidity	Cool 4 deg C	LIRO01	N11	UDS-4-26-72319	07/23/2019	SM2130 B
07/25/2019	Water	K3972-01	Turbidity	Cool 4 deg C	LIRO01	N11	UDS-4-40-72319	07/23/2019	SM2130 B

Date/Time 07-24-19 9:30  
 Received by: [Signature]  
 Relinquished by: [Signature]

Date/Time 07-24-19 12:15  
 Received by: [Signature]  
 Relinquished by: [Signature]