



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: Apatel

Run Number: LB104192

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/24/2019	12:20
2	K3891-07		1	8	100.00	>212.0	>212.0	07/24/2019	12:45
3	K3891-07DUP		1	8	100.00	>212.0	>212.0	07/24/2019	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 104192

WorkList Name : FP72419-1 WorkList ID : 128968 Department : Wet-Chemistry Date : 07-24-2019 08:55:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/30/2019	Solid	K3891-07	Flash Point	Cool 4 deg C	ROYF02	M63	P001-SW03-01	07/16/2019	1010A

Date/Time 07/24/19 11:45
 Received by: KS
 Relinquished by: CP

Date/Time 07/24/19 13:25
 Received by: CP
 Relinquished by: KS