



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : AMANDEEP

Run Number: LB104200

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/24/2019	17:15
2	CAL2	1	Water	NA	NA	20.2	7.00	07/24/2019	17:16
3	CAL3	1	Water	NA	NA	20.2	10.03	07/24/2019	17:17
4	ICV	1	Water	NA	NA	20.2	7.01	07/24/2019	17:19
5	CCV1	1	Water	NA	NA	20.2	2.02	07/24/2019	17:20
6	K3614-02	1	Solid	20.02	20	21.9	7.93	07/24/2019	17:22
7	K3614-02DUP	1	Solid	20.03	20	21.8	7.94	07/24/2019	17:24
8	K3961-04	1	Solid	20.05	20	21.5	7.44	07/24/2019	17:25
9	CCV2	1	Water	NA	NA	20.2	12.02	07/24/2019	17:26

VB 104200

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K3614 WorkList ID : 128990 Department : Wet-Chemistry Date : 07-24-2019 17:04:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2019	Solid	K3614-02	Corrosivity	Cool 4 deg C	PSEG05	M61	BU-03-072419	07/24/2019	9045D
07/23/2019	Solid	K3961-04	Corrosivity	Cool 4 deg C	PSEG01	M61	TP-10-A	07/23/2019	9045D

Date/Time 07-24-19 17:10
Received by: Jn
Relinquished by: CQ

Date/Time 07-25-19 18:00
Received by: [Signature]
Relinquished by: Jb