


ISM02.4

LB104207

INSTRUMENT ID: CU1



A= 0.0000e+000

B= 2.2662e-004

C= 2.0695e-002

Rho= 0.9997691

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.032	0.032	48	0.000	48					—
0.05	0.050										
0.20	0.200	0.204	0.004	808	0.0 %	808					+2
2.50	2.500	2.529	0.029	11067	0.0 %	11067					1
5.00	5.000	4.967	-0.033	21828	0.0 %	21828					-1
7.50	7.500	7.353	-0.147	32357	0.0 %	32357					-2
10.0	10.000	10.115	0.115	44545	0.0 %	44545					1

LB104207
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	48	-		0 ISM02.4	PPB	7/25/2019 10:21	Std
	0.2 S0.2	808	-		0.2 ISM02.4	PPB	7/25/2019 10:24	Std
	2.5 S2.5	11067	-		2.5 ISM02.4	PPB	7/25/2019 10:26	Std
	5 S5.0	21828	-		5 ISM02.4	PPB	7/25/2019 10:28	Std
	7.5 S7.5	32357	-		7.5 ISM02.4	PPB	7/25/2019 10:31	Std
	10 S10.0	44545	-		10 ISM02.4	PPB	7/25/2019 10:33	Std
ICV	ICV	17882	4.0731 -		ISM02.4	PPB	7/25/2019 10:36	SMPL
ICB	ICB	3	0.0214 -		ISM02.4	PPB	7/25/2019 10:38	SMPL
CCV085	CCV085	21794	4.9596 -		ISM02.4	PPB	7/25/2019 10:40	SMPL
CCB085	CCB085	181	0.0617 -		ISM02.4	PPB	7/25/2019 10:42	SMPL
PB121710BL	PBS710	46	0.0311 -		ISM02.4	PPB	7/25/2019 10:45	SMPL
K3821-01	MA52G9	4869	1.1241 -		ISM02.4	PPB	7/25/2019 10:47	SMPL
K3821-02	MA52H7	415	0.1147 -		ISM02.4	PPB	7/25/2019 10:49	SMPL
K3821-03	MA52H8	399	0.1111 -		ISM02.4	PPB	7/25/2019 10:51	SMPL
K3821-04	MA52J1	794	0.2006 -		ISM02.4	PPB	7/25/2019 10:54	SMPL
K3821-06	MA52T3	27399	6.2298 -		ISM02.4	PPB	7/25/2019 10:56	SMPL
K3821-07	MA52G0	156	0.056 -		ISM02.4	PPB	7/25/2019 10:58	SMPL
K3821-08	MA52G7	1007	0.2489 -		ISM02.4	PPB	7/25/2019 11:01	SMPL
K3821-09	MA52G7D	1027	0.2534 -		ISM02.4	PPB	7/25/2019 11:03	SMPL
K3821-10	MA52G7S	14946	3.4077 -		ISM02.4	PPB	7/25/2019 11:05	SMPL
PB121711BL	PBS711	24	0.0261 -		ISM02.4	PPB	7/25/2019 11:07	SMPL
K3851-01	MBENP0	9180	2.101 -		ISM02.4	PPB	7/25/2019 11:10	SMPL
K3851-02	MBENP1	11595	2.6483 -		ISM02.4	PPB	7/25/2019 11:12	SMPL
K3851-03	MBENP1D	11523	2.632 -		ISM02.4	PPB	7/25/2019 11:14	SMPL
K3851-04	MBENP1S	25376	5.7713 -		ISM02.4	PPB	7/25/2019 11:16	SMPL
K3851-05	MBENP4	20326	4.6269 -		ISM02.4	PPB	7/25/2019 11:19	SMPL
K3851-06	MBENP5	45449	10.3202 -		ISM02.4	PPB	7/25/2019 11:21	SMPL
K3851-07	MBENQ0	2954	0.6901 -		ISM02.4	PPB	7/25/2019 11:23	SMPL
K3851-08	MBENQ1	4370	1.011 -		ISM02.4	PPB	7/25/2019 11:26	SMPL
K3851-09	MBENQ3	4966	1.1461 -		ISM02.4	PPB	7/25/2019 11:28	SMPL
CCV086	CCV086	21743	4.948 -		ISM02.4	PPB	7/25/2019 11:31	SMPL
CCB086	CCB086	-12	0.018 -		ISM02.4	PPB	7/25/2019 11:33	SMPL
K3851-10	MBENR6	9242	2.1151 -		ISM02.4	PPB	7/25/2019 11:35	SMPL
K3851-11	MBENR7	16640	3.7916 -		ISM02.4	PPB	7/25/2019 11:37	SMPL
K3851-12	MBENR9	22410	5.0992 -		ISM02.4	PPB	7/25/2019 11:40	SMPL
K3851-13	MBENS0	7027	1.6131 -		ISM02.4	PPB	7/25/2019 11:42	SMPL
K3851-14	MBENS1	6214	1.4289 -		ISM02.4	PPB	7/25/2019 11:44	SMPL
K3851-15	MBENS2	34645	7.8718 -		ISM02.4	PPB	7/25/2019 11:46	SMPL
K3851-16	MBENR8	35435	8.0509 -		ISM02.4	PPB	7/25/2019 11:49	SMPL
K3851-17	MBENS3	3618	0.8406 -		ISM02.4	PPB	7/25/2019 11:51	SMPL
K3851-18	MBENS4	3439	0.8 -		ISM02.4	PPB	7/25/2019 11:53	SMPL
K3851-19	MBENS5	88921	20.1717 -		ISM02.4	PPB	7/25/2019 11:56	SMPL
K3851-20	MBENS6	22340	5.0833 -		ISM02.4	PPB	7/25/2019 11:58	SMPL
K3851-21	MBENS7	312	0.0914 -		ISM02.4	PPB	7/25/2019 12:00	SMPL
PB121712BL	PBS712	28	0.027 -		ISM02.4	PPB	7/25/2019 12:02	SMPL
K3895-01	MBENY7	7965	1.8257 -		ISM02.4	PPB	7/25/2019 12:05	SMPL

INSTRUMENT ID : CV1

K3895-02	MBEP00	7707	1.7672 -	ISM02.4	PPB	7/25/2019 12:07 SMPL
K3895-03	MBEP01	6763	1.5533 -	ISM02.4	PPB	7/25/2019 12:09 SMPL
K3895-04	MBEP02	1279	0.3105 -	ISM02.4	PPB	7/25/2019 12:12 SMPL
K3895-05	MBEP11	16427	3.7433 -	ISM02.4	PPB	7/25/2019 12:14 SMPL
K3895-06	MBEP11D	16472	3.7535 -	ISM02.4	PPB	7/25/2019 12:16 SMPL
K3895-07	MBEP11S	23900	5.4368 -	ISM02.4	PPB	7/25/2019 12:18 SMPL
CCV087	CCV087	21780	4.9564 -	ISM02.4	PPB	7/25/2019 12:21 SMPL
CCB087	CCB087	6	0.0221 -	ISM02.4	PPB	7/25/2019 12:23 SMPL
K3895-08	MBEP12	14163	3.2303 -	ISM02.4	PPB	7/25/2019 12:25 SMPL
K3895-09	MBEP13	10367	2.37 -	ISM02.4	PPB	7/25/2019 12:28 SMPL
K3895-10	MBEP14	21847	4.9716 -	ISM02.4	PPB	7/25/2019 12:30 SMPL
K3895-11	MBEP15	45348	10.2973 -	ISM02.4	PPB	7/25/2019 12:32 SMPL
K3895-12	MBEP16	62094	14.0922 -	ISM02.4	PPB	7/25/2019 12:34 SMPL
K3895-13	MBEP17	52108	11.8292 -	ISM02.4	PPB	7/25/2019 12:37 SMPL
K3895-14	MBEP18	61375	13.9293 -	ISM02.4	PPB	7/25/2019 12:39 SMPL
K3895-15	MBEP19	8206	1.8803 -	ISM02.4	PPB	7/25/2019 12:41 SMPL
K3895-16	MBEP20	559	0.1474 -	ISM02.4	PPB	7/25/2019 12:43 SMPL
K3895-17	MBEP21	38009	8.6342 -	ISM02.4	PPB	7/25/2019 12:46 SMPL
K3895-18	MBEP22	9735	2.2268 -	ISM02.4	PPB	7/25/2019 12:48 SMPL
K3895-19	MBEP23	13057	2.9796 -	ISM02.4	PPB	7/25/2019 12:50 SMPL
K3895-20	MBEP24	1975	0.4683 -	ISM02.4	PPB	7/25/2019 12:53 SMPL
K3895-21	MBEP32	614	0.1598 -	ISM02.4	PPB	7/25/2019 12:55 SMPL
K3851-06DLX2		21792	4.9591 -	ISM02.4	PPB	7/25/2019 12:57 SMPL
K3851-19DLX5		17109	3.8979 -	ISM02.4	PPB	7/25/2019 12:59 SMPL
K3895-11DLX2		22625	5.1479 -	ISM02.4	PPB	7/25/2019 13:02 SMPL
K3895-12DLX5		12288	2.8054 -	ISM02.4	PPB	7/25/2019 13:04 SMPL
K3895-13DLX2		25792	5.8656 -	ISM02.4	PPB	7/25/2019 13:06 SMPL
K3895-14DLX5		11960	2.731 -	ISM02.4	PPB	7/25/2019 13:09 SMPL
CCV088	CCV088	21559	4.9063 -	ISM02.4	PPB	7/25/2019 13:11 SMPL
CCB088	CCB088	4	0.0216 -	ISM02.4	PPB	7/25/2019 13:13 SMPL