

LB104217


7471B

INSTRUMENT ID: CU1

Linear

 μ Abs.:

54788



A= 0.0000e+000

B= 1.8501e-004

C= 3.3484e-002

Rho= 0.9995044

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.047	0.047	74	0.000	74				
0.2	0.200	0.218	0.018	997	0.0 %	997				
2.5	2.500	2.560	0.060	13656	0.0 %	13656				
5.0	5.000	4.888	-0.112	26241	0.0 %	26241				
7.5	7.500	7.317	-0.183	39368	0.0 %	39368				
10.0	10.000	10.170	0.170	54788	0.0 %	54788				

LB104217 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Jul 2019 14:21:38

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50-0	74	-	0.00007471B	PPB	25 Jul 2019 14:23:13	S	Std
0.2 - 1	50-2	997	-	0.20007471B	PPB	25 Jul 2019 14:25:32	S	Std
2.5 - 1	52-5	13656	-	2.50007471B	PPB	25 Jul 2019 14:27:50	S	Std
5.0 - 1	55-0	26241	-	5.00007471B	PPB	25 Jul 2019 14:30:08	S	Std
7.5 - 1	57-5	39368	-	7.50007471B	PPB	25 Jul 2019 14:38:25	S	Std
10.0 - 1	510-0	54788	-	10.00007471B	PPB	25 Jul 2019 14:40:42	S	Std
ICV47 - 1	ICV47	21862	4.0781	-7471B	PPB	25 Jul 2019 14:44:44	U	SMPL
ICB47 - 1	ICB47	7	0.0348	-7471B	PPB	25 Jul 2019 14:47:00	U	SMPL
CCV89 - 1	CCV89	26355	4.9094	-7471B	PPB	25 Jul 2019 14:49:16	U	SMPL
CCB89 - 1	CCB89	-18	0.0302	-7471B	PPB	25 Jul 2019 14:51:32	U	SMPL
CRA - 1	CRA	978	0.2144	-7471B	PPB	25 Jul 2019 14:53:49	U	SMPL
HighStd - 1	HighStd	54720	10.1572	-7471B	PPB	25 Jul 2019 14:56:05	U	SMPL
ChkStd - 1	ChkStd	36783	6.8387	-7471B	PPB	25 Jul 2019 14:58:21	U	SMPL
PB121708BL - 1	PBS	-25	0.0289	-7471B	PPB	25 Jul 2019 15:00:40	U	SMPL
PB121708BS - 1	LCSS	20281	3.7856	-7471B	PPB	25 Jul 2019 15:02:58	U	SMPL
K3614-01 - 1	BU-03-072419	1223	0.2597	-7471B	PPB	25 Jul 2019 15:05:15	U	SMPL
K3614-01DUP - 1	BU-03-072419DUP	1253	0.2653	-7471B	PPB	25 Jul 2019 15:07:33	U	SMPL
K3614-01MS - 1	BU-03-072419MS	21264	3.9675	-7471B	PPB	25 Jul 2019 15:09:50	U	SMPL
K3614-01MSD - 1	BU-03-072419MSD	21603	4.0302	-7471B	PPB	25 Jul 2019 15:12:08	U	SMPL
K3961-03 - 1	TP-10-A	8749	1.6521	-7471B	PPB	25 Jul 2019 15:14:24	U	SMPL
CCV90 - 1	CCV90	26335	4.9057	-7471B	PPB	25 Jul 2019 15:16:40	U	SMPL
CCB90 - 1	CCB90	-28	0.0283	-7471B	PPB	25 Jul 2019 15:18:56	U	SMPL
K3983-01 - 1	SP-1	25406	4.7338	-7471B	PPB	25 Jul 2019 15:21:13	U	SMPL
K3983-02 - 1	SP-2	31150	5.7965	-7471B	PPB	25 Jul 2019 15:23:29	U	SMPL
K3983-03 - 1	SP-3	22528	4.2014	-7471B	PPB	25 Jul 2019 15:25:46	U	SMPL
K3983-04 - 1	SP-4	23366	4.3564	-7471B	PPB	25 Jul 2019 15:28:03	U	SMPL
K3983-05 - 1	SP-5	23317	4.3473	-7471B	PPB	25 Jul 2019 15:30:20	U	SMPL
K3990-01 - 1	TP-1-A	6959	1.3210	-7471B	PPB	25 Jul 2019 15:32:38	U	SMPL
K3990-02 - 1	TP-1-B	1486	0.3084	-7471B	PPB	25 Jul 2019 15:34:55	U	SMPL
K3990-03 - 1	TP-2-A	5901	1.1252	-7471B	PPB	25 Jul 2019 15:37:13	U	SMPL
K3990-04 - 1	TP-2-B	4330	0.8346	-7471B	PPB	25 Jul 2019 15:39:30	U	SMPL
K3990-05 - 1	TP-3-A	311	0.0910	-7471B	PPB	25 Jul 2019 15:41:47	U	SMPL
CCV91 - 1	CCV91	26494	4.9351	-7471B	PPB	25 Jul 2019 15:44:03	U	SMPL
CCB91 - 1	CCB91	-19	0.0300	-7471B	PPB	25 Jul 2019 15:46:20	U	SMPL
K3990-06 - 1	TP-3-B	2956	0.5804	-7471B	PPB	25 Jul 2019 15:48:36	U	SMPL
K3990-07 - 1	TP-4-A	3930	0.7606	-7471B	PPB	25 Jul 2019 15:50:52	U	SMPL
K3990-08 - 1	TP-4-B	3568	0.6936	-7471B	PPB	25 Jul 2019 15:53:08	U	SMPL
K3990-09 - 1	TP-5-A	11507	2.1624	-7471B	PPB	25 Jul 2019 15:55:25	U	SMPL
K3990-10 - 1	TP-5-B	3033	0.5946	-7471B	PPB	25 Jul 2019 15:57:41	U	SMPL
PB121709BL - 1	PBS	31	0.0392	-7471B	PPB	25 Jul 2019 16:00:00	U	SMPL
PB121709BS - 1	LCSS	20248	3.7795	-7471B	PPB	25 Jul 2019 16:02:19	U	SMPL
K3591-01 - 1	LOD-MDL-SOIL-01-QT3-2019	155	0.0622	-7471B	PPB	25 Jul 2019 16:23:22	U	SMPL
K3591-02 - 1	LOQ-SOIL-02-QT3-2019	2169	0.4348	-7471B	PPB	25 Jul 2019 16:25:40	U	SMPL
K3614-01LX5 - 1		274	0.0842	-7471B	PPB	25 Jul 2019 16:27:56	U	SMPL
CCV92 - 1	CCV92	26087	4.8598	-7471B	PPB	25 Jul 2019 16:30:13	U	SMPL
CCB92 - 1	CCB92	92	0.0505	-7471B	PPB	25 Jul 2019 16:32:29	U	SMPL
K3614-01A - 1		21658	4.0404	-7471B	PPB	25 Jul 2019 16:34:45	U	SMPL
CCV93 - 1	CCV93	26014	4.8463	-7471B	PPB	25 Jul 2019 16:37:02	U	SMPL
CCB93 - 1	CCB93	-59	0.0226	-7471B	PPB	25 Jul 2019 16:39:19	U	SMPL