



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 7/26/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 07/25/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 08:20
Out Date: 07/26/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104228

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K3620-11	CL-02-072519-A	1	1.14	9.61	8.91	91.7
K3620-13	CL-02-072519-B	2	1.17	9.66	8.85	90.5
K3620-15	CL-02-072519-C	3	1.18	9.9	9.18	91.7
K3620-17	CL-02-072519-D	4	1.18	9.75	9.06	91.9
K3620-19	CL-02-072519-E	5	1.11	9.95	9.22	91.7
K3626-09	SU-01-072519-A	6	1.14	9.52	8.54	88.3
K3626-11	SU-01-072519-B	7	1.15	9.97	9.21	91.4
K3626-13	SU-03-072519-A	8	1.17	9.78	9.17	92.9
K3626-15	SU-03-072519-B	9	1.18	9.67	8.92	91.2
K4000-01	COMP-1	10	1.00	2.00	2.00	100.0
K4000-02	COMP-2	11	1.00	2.00	2.00	100.0
K4000-03	COMP-3	12	1.00	2.00	2.00	100.0
K4000-04	COMP-4	13	1.00	2.00	2.00	100.0
K4000-05	COMP-5	14	1.00	2.00	2.00	100.0
K4000-06	COMP-6	15	1.00	2.00	2.00	100.0
K4000-07	COMP-7	16	1.00	2.00	2.00	100.0
K4005-01	S13-03-B	17	1.14	9.51	8.52	88.2
K4005-02	S13-03-S	18	1.15	9.75	8.22	82.2
K4005-03	S13-04-B	19	1.14	9.97	8.27	80.7
K4005-04	S13-04-S	20	1.14	9.55	7.91	80.5
K4005-05	DUP-11	21	1.13	9.71	8.82	89.6
K4005-06	S14-01-S	22	1.15	9.71	8.66	87.7
K4005-07	S14-01-B	23	1.13	9.94	9.1	90.5
K4005-08	S14-02-S	24	1.13	9.86	8.99	90.0
K4006-01	S8-10-B	25	1.13	9.71	8.13	81.6
K4006-02	S8-11-B	26	1.14	9.98	8.65	85.0
K4006-03	S8-12-B	27	1.15	9.96	8.59	84.4
K4006-04	S8-13-B	28	1.14	9.85	8.49	84.4
K4006-05	S8-14-B	29	1.14	9.63	8.53	87.0
K4006-06	S5-08-B	30	1.13	9.85	9.14	91.9
K4006-07	S6-01-B	31	1.13	9.66	9.00	92.3
K4006-08	S6-01-S	32	1.12	9.73	9.16	93.4
K4006-09	S6-02-B	33	1.15	9.5	8.59	89.1
K4006-10	S6-03-B	34	1.14	9.97	8.99	88.9
K4006-11	S6-04-B	35	1.13	9.75	8.83	89.3
K4006-12	DUP-6	36	1.13	9.98	8.92	88.0
K4006-13	S10-01-S-01	37	1.14	9.87	8.87	88.5
K4011-01	DRUM-COMP-OS	38	1.12	9.78	9.16	92.8
K4011-02	CONCRETE-SLAB	39	1.00	2.00	2.00	100.0



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 7/26/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 07/25/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 08:20
Out Date: 07/26/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104228

<u>Lab ID</u>	<u>Client Sample ID</u>	<u>Dish#</u>	<u>Dish Wt(g)</u> <u>(A)</u>	<u>Dish +</u> <u>Sample Wt(g)</u> <u>(B)</u>	<u>Dish + Dry</u> <u>Sample Wt(g)</u> <u>(C)</u>	<u>% Solid</u>
K4011-03	CONCRETE-PIPE	40	1.00	2.00	2.00	100.0
K4012-02	OILY-DEBRIS	41	1.00	2.00	2.00	100.0
K4012-05	OIL	42	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$