



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104229

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/25/2019	18:00
2	CAL2	1	Water	NA	NA	20.2	7.01	07/25/2019	18:01
3	CAL3	1	Water	NA	NA	20.2	10.02	07/25/2019	18:02
4	ICV	1	Water	NA	NA	20.2	6.99	07/25/2019	18:04
5	CCV1	1	Water	NA	NA	20.2	2.02	07/25/2019	18:05
6	K3620-12	1	Solid	20.02	20	23.4	8.44	07/25/2019	18:06
7	K3620-14	1	Solid	20.03	20	22.8	8.11	07/25/2019	18:08
8	K3620-16	1	Solid	20.04	20	23.2	8.09	07/25/2019	18:10
9	K3620-18	1	Solid	20.02	20	23.3	7.72	07/25/2019	18:11
10	K3620-20	1	Solid	20.01	20	23.8	9.44	07/25/2019	18:14
11	K3626-10	1	Solid	20.02	20	22.4	8.02	07/25/2019	18:15
12	K3626-12	1	Solid	20.04	20	22.6	8.59	07/25/2019	18:16
13	K3626-14	1	Solid	20.03	20	22.2	9.40	07/25/2019	18:18
14	K3626-16	1	Solid	20.02	20	22.4	8.72	07/25/2019	18:20
15	K3626-16DUP	1	Solid	20.01	20	22.5	8.73	07/25/2019	18:21
16	CCV2	1	Water	NA	NA	20.2	12.02	07/25/2019	18:24
17	K4012-01	1	Solid	20.01	20	22.3	6.08	07/25/2019	18:25
18	K4012-04	1	Solid	20.03	20	23.0	5.29	07/25/2019	18:26
19	CCV3	1	Water	NA	NA	20.1	2.02	07/25/2019	18:27

20104229

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K3629 WorkList ID : 129047 Department : Wet-Chemistry Date : 07-25-2019 15:04:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/25/2019	Solid	K3620-12-13	Corrosivity	Cool 4 deg C	PSEG05		CL-02-072519-A	07/25/2019	9045D
07/25/2019	Solid	K3620-14-13	Corrosivity	Cool 4 deg C	PSEG05		CL-02-072519-B	07/25/2019	9045D
07/25/2019	Solid	K3620-16-13	Corrosivity	Cool 4 deg C	PSEG05		CL-02-072519-C	07/25/2019	9045D
07/25/2019	Solid	K3620-18-13	Corrosivity	Cool 4 deg C	PSEG05		CL-02-072519-D	07/25/2019	9045D
07/25/2019	Solid	K3620-20-13	Corrosivity	Cool 4 deg C	PSEG05		CL-02-072519-E	07/25/2019	9045D
07/25/2019	Solid	K3626-10-13	Corrosivity	Cool 4 deg C	PSEG05		SU-01-072519-A	07/25/2019	9045D
07/25/2019	Solid	K3626-12-13	Corrosivity	Cool 4 deg C	PSEG05		SU-01-072519-B	07/25/2019	9045D
07/25/2019	Solid	K3626-14-13	Corrosivity	Cool 4 deg C	PSEG05		SU-03-072519-A	07/25/2019	9045D
07/25/2019	Solid	K3626-16-13	Corrosivity	Cool 4 deg C	PSEG05		SU-03-072519-B	07/25/2019	9045D
07/25/2019	Solid	K4012-01-13	Corrosivity	Cool 4 deg C	PSEG01	M62	OILY-DEBRIS	07/25/2019	9045D
07/25/2019	Solid	K4012-04-13	Corrosivity	Cool 4 deg C	PSEG01	M62	OIL	07/25/2019	9045D

Date/Time 07-25-19 15:15
 Received by: JH
 Relinquished by: JH

Date/Time 07-25-19 18:15
 Received by: JH
 Relinquished by: JH