



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104234

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/25/2019	16:30
2	CAL2	1	Water	NA	NA	20.2	7.00	07/25/2019	16:31
3	CAL3	1	Water	NA	NA	20.2	10.04	07/25/2019	16:32
4	ICV	1	Water	NA	NA	20.2	6.99	07/25/2019	16:33
5	CCV1	1	Water	NA	NA	20.1	2.02	07/25/2019	16:35
6	K3995-01	1	Water	NA	NA	20.5	7.86	07/25/2019	16:36
7	K3995-01DUP	1	Water	NA	NA	20.6	7.87	07/25/2019	16:37
8	CCV2	1	Water	NA	NA	20.2	12.02	07/25/2019	16:40

LB104234

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH K3995 WorkList ID: 129063 Department: Wet-Chemistry Date: 07-25-2019 16:15:354

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/25/2019	Water	K3995-01	pH	Cool 4 deg C	PSEG01	M61	A1014	07/25/2019	9040C

Date/Time 07-25-14 16:20
Received by: JT
Relinquished by: JT

Date/Time 07-25-14 17:00
Received by: JT
Relinquished by: JT