



## Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB104237

| Reagent/Standard                       | Lot/Log # |
|----------------------------------------|-----------|
| Residual chlorine LOD, 0.05PPM         | WP76810   |
| Residual chlorine ICV-LCS, 0.4PPM      | WP76809   |
| Chlorine Calibration std, 0.1ppm       | WP76804   |
| Chlorine Calibration std, 0.2ppm       | WP76805   |
| Chlorine Calibration std, 0.8ppm       | WP76807   |
| Chlorine Calibration std, 0.0ppm       | WP76803   |
| Chlorine Calibration std, 1.6ppm       | WP76808   |
| Chlorine Calibration std & CCV, 0.4ppm | WP76806   |
| Total Chlorine Powder Pillows          | W2249     |

Intercept: 0.0000

Slope: 1.0000

Regression: 1.000000

| Seq | Lab ID | True Val<br>(mg/l) | Dilution | Initial<br>Reading | Final<br>Reading | Difference | AnalDate   | Anal Time |
|-----|--------|--------------------|----------|--------------------|------------------|------------|------------|-----------|
| 1   | CAL1   | 0                  | 1        | 0.000              | 0.000            | 0.000      | 07/26/2019 | 10:13     |
| 2   | CAL2   | 0.1                | 1        | 0.000              | 0.100            | 0.100      | 07/26/2019 | 10:14     |
| 3   | CAL3   | 0.2                | 1        | 0.000              | 0.200            | 0.200      | 07/26/2019 | 10:15     |
| 4   | CAL4   | 0.4                | 1        | 0.000              | 0.400            | 0.400      | 07/26/2019 | 10:16     |
| 5   | CAL5   | 0.8                | 1        | 0.000              | 0.800            | 0.800      | 07/26/2019 | 10:17     |
| 6   | CAL6   | 1.6                | 1        | 0.000              | 1.600            | 1.600      | 07/26/2019 | 10:18     |



## Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB104237

| Seq | Lab ID     | Initial Weight | Final Vol | True Value (mg/L) | DF | Initial Reading | Final Reading | Diff.  | Result (mg/L) | Anal Date  | Anal Time |
|-----|------------|----------------|-----------|-------------------|----|-----------------|---------------|--------|---------------|------------|-----------|
| 1   | ICV        |                |           | 0.4               | 1  | 0.0000          | 0.3800        | 0.3800 | 0.3800        | 07/26/2019 | 10:19     |
| 2   | ICB        |                |           |                   | 1  | 0.0000          | 0.0000        | 0.0000 | 0.0000        | 07/26/2019 | 10:21     |
| 3   | CCV1       |                |           | 0.4               | 1  | 0.0000          | 0.3900        | 0.3900 | 0.3900        | 07/26/2019 | 10:22     |
| 4   | CCB1       |                |           |                   | 1  | 0.0000          | 0.0000        | 0.0000 | 0.0000        | 07/26/2019 | 10:22     |
| 5   | LB104237BL | 50             | 50        |                   | 1  | 0.0000          | 0.0000        | 0.0000 | 0.0000        | 07/26/2019 | 10:23     |
| 6   | LB104237BS | 50             | 50        | 0.4               | 1  | 0.0000          | 0.3600        | 0.3600 | 0.3600        | 07/26/2019 | 10:24     |
| 7   | K3591-09   | 50             | 50        |                   | 1  | 0.0000          | 0.0480        | 0.0480 | 0.0480        | 07/26/2019 | 10:25     |
| 8   | CCV2       |                |           | 0.4               | 1  | 0.0000          | 0.4100        | 0.4100 | 0.4100        | 07/26/2019 | 10:26     |
| 9   | CCB2       |                |           |                   | 1  | 0.0000          | 0.0000        | 0.0000 | 0.0000        | 07/26/2019 | 10:27     |