

LB104239

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : CT

Instrument ID : Konelab

7/26/2019 10:34


Test Phenolics-

Accepted 7/26/2019 10:34

Factor 9.124
Bias 0.012

Coeff. of det. 0.999659

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.017	0.0457	0.0500	
2	Phenol-2	0.024	0.1126	0.1000	
3	Phenol-2	0.039	0.2511	0.2500	
4	Phenol-2	0.067	0.5052	0.5000	
5	Phenol-2	0.119	0.9748	1.0000	
6	Phenol-2	0.232	2.0106	2.0000	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : CT

Instrument ID : Konelab

7/26/2019 11:15

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.000	0.0	0.121	
ICB1	0.004	0.0	0.012	
CCV1	0.985	0.0	0.120	
CCB1	0.008	0.0	0.013	
PB121730BL	0.002	0.0	0.012	
PB121730BS	1.011	0.0	0.123	
K3992-01	0.008	0.0	0.013	
K3992-01DUP	0.009	0.0	0.013	
K3992-01MSD	0.946	0.0	0.116	
K3992-01MS	0.949	0.0	0.116	
CCV2	0.978	0.0	0.119	
CCB2	-0.001	0.0	0.012	

N	12
Mean	0.491
SD	0.5085
CV%	103.46

Aquakem v. 7.2AQ1

Results from time period:

Fri Jul 26 10:33:41 2019

Fri Jul 26 11:13:23 2019

CT

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0457	mg/l	7/26/2019 10:33:41	
0.1PPM	A	Phenolics-	P	0.1126	mg/l	7/26/2019 10:33:42	
0.25PPM	A	Phenolics-	P	0.2511	mg/l	7/26/2019 10:33:43	
0.50PPM	A	Phenolics-	P	0.5052	mg/l	7/26/2019 10:33:44	
1.0PPM	A	Phenolics-	P	0.9748	mg/l	7/26/2019 10:33:45	
2.0PPM	A	Phenolics-	P	2.0106	mg/l	7/26/2019 10:33:46	
ICV1	S	Phenolics-	P	0.9995	mg/l	7/26/2019 11:09:24	
ICB1	S	Phenolics-	P	0.004	mg/l	7/26/2019 11:09:25	
CCV1	S	Phenolics-	P	0.9848	mg/l	7/26/2019 11:09:26	
CCB1	S	Phenolics-	P	0.0077	mg/l	7/26/2019 11:09:27	
PB121730BL	S	Phenolics-	P	0.0018	mg/l	7/26/2019 11:09:28	
PB121730BS	S	Phenolics-	P	1.0105	mg/l	7/26/2019 11:09:29	
K3992-01	S	Phenolics-	P	0.0076	mg/l	7/26/2019 11:09:30	
K3992-01DUP	S	Phenolics-	P	0.0093	mg/l	7/26/2019 11:09:31	
K3992-01MSD	S	Phenolics-	P	0.9463	mg/l	7/26/2019 11:09:34	
K3992-01MS	S	Phenolics-	P	0.9489	mg/l	7/26/2019 11:09:35	
CCV2	S	Phenolics-	P	0.9778	mg/l	7/26/2019 11:13:22	
CCB2	S	Phenolics-	P	-0.0007	mg/l	7/26/2019 11:13:23	