


LB104241

7470A

INSTRUMENT ID : CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	54	0.000	54				
0.2	0.200	0.187	-0.013	708	0.0 %	708				
2.5	2.500	2.532	0.032	9833	0.0 %	9833				
5.0	5.000	4.877	-0.123	18958	0.0 %	18958				
7.5	7.500	7.638	0.138	29701	0.0 %	29701				
10.0	10.000	9.946	-0.054	38683	0.0 %	38683				

LB104241 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Jul 2019 13:01:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	54	-	0.0000	7470A	PPB	26 Jul 2019 13:13:59	S	Std
0.2 - 1	50-2	708	-	0.2000	7470A	PPB	26 Jul 2019 13:16:09	S	Std
2.5 - 1	52-5	9833	-	2.5000	7470A	PPB	26 Jul 2019 13:18:17	S	Std
5.0 - 1	55-0	18958	-	5.0000	7470A	PPB	26 Jul 2019 13:20:27	S	Std
7.5 - 1	57-5	29701	-	7.5000	7470A	PPB	26 Jul 2019 13:22:36	S	Std
10.0 - 1	510-0	38683	-	10.0000	7470A	PPB	26 Jul 2019 13:27:42	S	Std
ICV48 - 1	ICV48	14938	3.8441	-	7470A	PPB	26 Jul 2019 13:30:11	U	SMPL
ICB48 - 1	ICB48	2	0.0058	-	7470A	PPB	26 Jul 2019 13:32:17	U	SMPL
CCV94 - 1	CCV94	18843	4.8476	-	7470A	PPB	26 Jul 2019 13:34:24	U	SMPL
CCB94 - 1	CCB94	10	0.0078	-	7470A	PPB	26 Jul 2019 13:36:30	U	SMPL
CRA - 1	CRA	720	0.1903	-	7470A	PPB	26 Jul 2019 13:38:36	U	SMPL
HighStd - 1	HighStd	38124	9.8026	-	7470A	PPB	26 Jul 2019 13:40:43	U	SMPL
ChkStd - 1	ChkStd	26821	6.8979	-	7470A	PPB	26 Jul 2019 13:42:49	U	SMPL
PB121751BL - 1	PBW	22	0.0109	-	7470A	PPB	26 Jul 2019 13:44:57	U	SMPL
PB121751BS - 1	LCSW	15265	3.9282	-	7470A	PPB	26 Jul 2019 13:47:05	U	SMPL
K3614-02 - 1	BU-03-072419	21	0.0107	-	7470A	PPB	26 Jul 2019 13:49:13	U	SMPL
K3614-02DUP - 1	BU-03-072419DUP	67	0.0225	-	7470A	PPB	26 Jul 2019 13:51:21	U	SMPL
K3614-02MS - 1	BU-03-072419MS	15675	4.0335	-	7470A	PPB	26 Jul 2019 13:53:29	U	SMPL
K3614-02MSD - 1	BU-03-072419MSD	16060	4.1325	-	7470A	PPB	26 Jul 2019 13:55:38	U	SMPL
K3620-12 - 1	CL-02-072519-A	38	0.0150	-	7470A	PPB	26 Jul 2019 13:57:45	U	SMPL
CCV95 - 1	CCV95	19066	4.9050	-	7470A	PPB	26 Jul 2019 13:59:52	U	SMPL
CCB95 - 1	CCB95	-3	0.0045	-	7470A	PPB	26 Jul 2019 14:01:59	U	SMPL
K3620-14 - 1	CL-02-072519-B	111	0.0338	-	7470A	PPB	26 Jul 2019 14:04:06	U	SMPL
K3620-16 - 1	CL-02-072519-C	88	0.0279	-	7470A	PPB	26 Jul 2019 14:06:12	U	SMPL
K3620-18 - 1	CL-02-072519-D	43	0.0163	-	7470A	PPB	26 Jul 2019 14:08:19	U	SMPL
K3620-20 - 1	CL-02-072519-E	53	0.0189	-	7470A	PPB	26 Jul 2019 14:10:25	U	SMPL
K3626-10 - 1	SU-01-072519-A	41	0.0158	-	7470A	PPB	26 Jul 2019 14:12:33	U	SMPL
K3626-12 - 1	SU-01-072519-B	75	0.0245	-	7470A	PPB	26 Jul 2019 14:14:40	U	SMPL
K3626-14 - 1	SU-03-072519-A	99	0.0307	-	7470A	PPB	26 Jul 2019 14:16:49	U	SMPL
K3626-16 - 1	SU-03-072519-B	78	0.0253	-	7470A	PPB	26 Jul 2019 14:18:57	U	SMPL
K3961-04 - 1	TP-10-A	69	0.0230	-	7470A	PPB	26 Jul 2019 14:21:06	U	SMPL
K4012-01 - 1	OILY-DEBRIS	64	0.0217	-	7470A	PPB	26 Jul 2019 14:23:13	U	SMPL
CCV96 - 1	CCV96	19062	4.9039	-	7470A	PPB	26 Jul 2019 14:25:19	U	SMPL
CCB96 - 1	CCB96	8	0.0073	-	7470A	PPB	26 Jul 2019 14:27:27	U	SMPL
K4012-04 - 1	OIL	456	0.1224	-	7470A	PPB	26 Jul 2019 14:29:34	U	SMPL
PB121751TB - 1	PB121722TB	80	0.0258	-	7470A	PPB	26 Jul 2019 14:31:42	U	SMPL
K3614-02LX5 - 1		33	0.0137	-	7470A	PPB	26 Jul 2019 14:33:49	U	SMPL
K3614-02A - 1		15913	4.0947	-	7470A	PPB	26 Jul 2019 14:36:00	U	SMPL
CCV97 - 1	CCV97	18970	4.8803	-	7470A	PPB	26 Jul 2019 14:38:08	U	SMPL
CCB97 - 1	CCB97	19	0.0101	-	7470A	PPB	26 Jul 2019 14:40:16	U	SMPL