



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104243

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/26/2019	12:15
2	CAL2	1	Water	NA	NA	20.1	6.99	07/26/2019	12:16
3	CAL3	1	Water	NA	NA	20.1	10.03	07/26/2019	12:17
4	ICV	1	Water	NA	NA	20.1	7.01	07/26/2019	12:19
5	CCV1	1	Water	NA	NA	20.1	2.02	07/26/2019	12:20
6	K4024-09	1	Water	NA	NA	20.3	7.01	07/26/2019	12:21
7	K4024-09DUP	1	Water	NA	NA	20.4	7.02	07/26/2019	12:24
8	CCV2	1	Water	NA	NA	20.1	12.02	07/26/2019	12:25

2810243 28104243

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH K4024 WorkList ID: 129078 Department: Wet-Chemistry Date: 07-26-2019 12:06:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/25/2019	Water	K4024-09-F	pH	Cool 4 deg C	PSEG01	M23	TP-8	07/25/2019	9040C

Date/Time 07-26-19 12:10
Received by: Jn
Relinquished by: CQ

Date/Time 07-26-19 13:15
Received by: CQ
Relinquished by: Jn