


LB104256

7471B

INSTRUMENT ID : CV1



Linear

A= 0.0000e+000

B= 1.9949e-004

C= 3.1246e-002

Rho= 0.9997180

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.046	0.046	75	0.000	75				
0.2	0.200	0.222	0.022	958	0.0 %	958				
2.5	2.500	2.544	0.044	12595	0.0 %	12595				
5.0	5.000	4.825	-0.175	24028	0.0 %	24028				
7.5	7.500	7.465	-0.035	37265	0.0 %	37265				
10.0	10.000	10.098	0.098	50461	0.0 %	50461				

LB104256 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jul 2019 15:07:02

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50-0	75	-	0.00007471B	PPB	26 Jul 2019 15:07:12	S	Std
0.2 - 1	50-2	958	-	0.20007471B	PPB	26 Jul 2019 15:09:30	S	Std
2.5 - 1	52-5	12595	-	2.50007471B	PPB	26 Jul 2019 15:11:48	S	Std
5.0 - 1	55-0	24028	-	5.00007471B	PPB	26 Jul 2019 15:14:05	S	Std
7.5 - 1	57-5	37265	-	7.50007471B	PPB	26 Jul 2019 15:16:23	S	Std
10.0 - 1	510-0	50461	-	10.00007471B	PPB	26 Jul 2019 15:23:01	S	Std
ICV49 - 1	ICV49	19431	3.9075	-7471B	PPB	26 Jul 2019 15:26:57	U	SMPL
ICB49 - 1	ICB49	-20	0.0273	-7471B	PPB	26 Jul 2019 15:29:13	U	SMPL
CCV98 - 1	CCV98	23945	4.8080	-7471B	PPB	26 Jul 2019 15:31:30	U	SMPL
CCB98 - 1	CCB98	-33	0.0247	-7471B	PPB	26 Jul 2019 15:33:46	U	SMPL
CRA - 1	CRA	891	0.2090	-7471B	PPB	26 Jul 2019 15:36:02	U	SMPL
HighStd - 1	HighStd	52588	10.5221	-7471B	PPB	26 Jul 2019 15:38:18	U	SMPL
ChkStd - 1	ChkStd	35545	7.1221	-7471B	PPB	26 Jul 2019 15:40:34	U	SMPL
PB121757BL - 1	PBS	-19	0.0275	-7471B	PPB	26 Jul 2019 15:42:53	U	SMPL
PB121757BS - 1	LCSS	19157	3.8529	-7471B	PPB	26 Jul 2019 15:45:13	U	SMPL
K3620-11 - 1	CL-02-072519-A	5355	1.0995	-7471B	PPB	26 Jul 2019 15:47:31	U	SMPL
K3620-11DUP - 1	CL-02-072519-ADUP	5509	1.1302	-7471B	PPB	26 Jul 2019 15:49:53	U	SMPL
K3620-11MS - 1	CL-02-072519-AMS	23888	4.7967	-7471B	PPB	26 Jul 2019 15:52:13	U	SMPL
K3620-11MSD - 1	CL-02-072519-AMSD	24167	4.8523	-7471B	PPB	26 Jul 2019 15:54:31	U	SMPL
K3620-13 - 1	CL-02-072519-B	4704	0.9696	-7471B	PPB	26 Jul 2019 15:56:48	U	SMPL
CCV99 - 1	CCV99	24316	4.8821	-7471B	PPB	26 Jul 2019 15:59:06	U	SMPL
CCB99 - 1	CCB99	-32	0.0249	-7471B	PPB	26 Jul 2019 16:01:22	U	SMPL
K3620-15 - 1	CL-02-072519-C	3264	0.6824	-7471B	PPB	26 Jul 2019 16:03:39	U	SMPL
K3620-17 - 1	CL-02-072519-D	16689	3.3605	-7471B	PPB	26 Jul 2019 16:05:56	U	SMPL
K3620-19 - 1	CL-02-072519-E	11668	2.3589	-7471B	PPB	26 Jul 2019 16:10:45	U	SMPL
K3626-09 - 1	SU-01-072519-A	3547	0.7388	-7471B	PPB	26 Jul 2019 16:13:01	U	SMPL
K3626-11 - 1	SU-01-072519-B	1410	0.3125	-7471B	PPB	26 Jul 2019 16:15:18	U	SMPL
K3626-13 - 1	SU-03-072519-A	1600	0.3504	-7471B	PPB	26 Jul 2019 16:17:35	U	SMPL
K3626-15 - 1	SU-03-072519-B	675	0.1659	-7471B	PPB	26 Jul 2019 16:19:52	U	SMPL
K4011-01 - 1	DRUM-COMP-OS	375	0.1061	-7471B	PPB	26 Jul 2019 16:22:10	U	SMPL
K4011-02 - 1	CONCRETE-SLAB	221	0.0753	-7471B	PPB	26 Jul 2019 16:24:28	U	SMPL
K4011-03 - 1	CONCRETE-PIPE	380	0.1071	-7471B	PPB	26 Jul 2019 16:26:44	U	SMPL
CCV01 - 1	CCV01	24605	4.9397	-7471B	PPB	26 Jul 2019 16:29:00	U	SMPL
CCB01 - 1	CCB01	-17	0.0279	-7471B	PPB	26 Jul 2019 16:31:16	U	SMPL
K4013-01 - 1	TANK-ABC-SLUDGE	3307	0.6910	-7471B	PPB	26 Jul 2019 16:33:32	U	SMPL
PB121758BL - 1	PBS	28	0.0368	-7471B	PPB	26 Jul 2019 16:35:52	U	SMPL
PB121758BS - 1	LCSS	19693	3.9598	-7471B	PPB	26 Jul 2019 16:38:08	U	SMPL
K3591-03 - 1	MDL-MDL-SOIL-03-QT3-2019	145	0.0602	-7471B	PPB	26 Jul 2019 16:40:24	U	SMPL
K3620-11LX5 - 1		1112	0.2531	-7471B	PPB	26 Jul 2019 16:42:42	U	SMPL
K3620-11A - 1		25037	5.0259	-7471B	PPB	26 Jul 2019 16:45:00	U	SMPL
CCV02 - 1	CCV02	24590	4.9367	-7471B	PPB	26 Jul 2019 16:47:19	U	SMPL
CCB02 - 1	CCB02	-15	0.0283	-7471B	PPB	26 Jul 2019 16:49:36	U	SMPL