



### Analytical Summary Report

Analysis Method: 9045D  
Parameter: pH  
Run Number: LB104258  
BalanceID: WC SC-4

Analyst By : JIGNESH  
Supervisor Review By : apatel  
Slope : 98.3  
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/26/2019	15:30
2	CAL2	1	Water	NA	NA	20.1	7.00	07/26/2019	15:31
3	CAL3	1	Water	NA	NA	20.1	10.03	07/26/2019	15:33
4	ICV	1	Water	NA	NA	20.2	6.99	07/26/2019	15:35
5	CCV1	1	Water	NA	NA	20.1	2.02	07/26/2019	15:36
6	K4034-01	1	Solid	20.02	20	21.9	7.38	07/26/2019	15:37
7	K4034-01DUP	1	Solid	20.03	20	21.8	7.39	07/26/2019	15:38
8	CCV2	1	Water	NA	NA	20.1	12.02	07/26/2019	15:40