



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104279

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/29/2019	10:50
2	K4012-04		1	9	100.00	>212.0	>212.0	07/29/2019	11:20
3	K4012-04DUP		1	9	100.00	>212.0	>212.0	07/29/2019	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP72919 WorkList ID : 129116 Department : Wet-Chemistry Date : 07-29-2019 09:17:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/01/2019	Solid	K4012-04	Flash Point	Cool 4 deg C	PSEG01	M62	OIL	07/25/2019	1010A

Date/Time 07/29/19 10:30
 Received by: KS
 Relinquished by: CP

Date/Time 07/29/19 11:50
 Received by: CP
 Relinquished by: KS