



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104280

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	07/29/2019	08:55
2	K3995-01		1	7	100.00	>212.0	>212.0	07/29/2019	09:20
3	K3995-01DUP		1	7	100.00	>212.0	>212.0	07/29/2019	09:45
4	K4024-10		1	8	100.00	>212.0	>212.0	07/29/2019	10:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 104280

WorkList Name : FP72619

WorkList ID : 129070

Department : Wet-Chemistry

Date : 07-26-2019 09:05:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/01/2019	Water	K3995-01	Flash Point	Cool 4 deg C	PSEG01	M61	A1014	07/25/2019	1010A
08/08/2019	Water	K4024-10	Flash Point	Cool 4 deg C	PSEG01	M23	TP-8	07/25/2019	1010A

Date/Time 07/29/19 08:30
 Received by: KS
 Relinquished by: KS

Date/Time 07/29/19 10:25
 Received by: KS
 Relinquished by: KS