


LB104302

ISM02.4

INSTRUMENT ID: C01

Linear



A= 0.0000e+000

B= 2.4281e-004

C= 6.9712e-003

Rho= 0.9999287

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.045	0.045	157	0.000	157					—
0.05	0.050										
0.20	0.200	0.219	0.019	875	0.0 %	875					
2.50	2.500	2.448	-0.052	10054	0.0 %	10054					
5.00	5.000	4.976	-0.024	20466	0.0 %	20466					
7.50	7.500	7.451	-0.049	30658	0.0 %	30658					
10.0	10.000	10.060	0.060	41403	0.0 %	41403					

10
-2
0
-1
1

LB104302
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	157	-		0 ISM02.4	PPB	7/30/2019 11:53	Std
	0.2 S0.2	875	-		0.2 ISM02.4	PPB	7/30/2019 11:56	Std
	2.5 S2.5	10054	-		2.5 ISM02.4	PPB	7/30/2019 11:58	Std
	5 S5.0	20466	-		5 ISM02.4	PPB	7/30/2019 12:00	Std
	7.5 S7.5	30658	-		7.5 ISM02.4	PPB	7/30/2019 12:03	Std
	10 S10.0	41403	-		10 ISM02.4	PPB	7/30/2019 12:05	Std
ICV	ICV	16475	4.0073 -		ISM02.4	PPB	7/30/2019 12:08	SMPL
ICB	ICB	113	0.0344 -		ISM02.4	PPB	7/30/2019 12:11	SMPL
CCV009	CCV009	20821	5.0625 -		ISM02.4	PPB	7/30/2019 12:13	SMPL
CCB009	CCB009	109	0.0334 -		ISM02.4	PPB	7/30/2019 12:15	SMPL
PB121828BL	PBW828	117	0.0354 -		ISM02.4	PPB	7/30/2019 12:17	SMPL
K3821-05	MA52S3	160	0.0458 -		ISM02.4	PPB	7/30/2019 12:20	SMPL
PB121829BL	PBW829	131	0.0388 -		ISM02.4	PPB	7/30/2019 12:22	SMPL
K3877-01	MA52V2	19299	4.6929 -		ISM02.4	PPB	7/30/2019 12:24	SMPL
K3877-02	MA52L2	85	0.0276 -		ISM02.4	PPB	7/30/2019 12:29	SMPL
K3877-03	MA52M6	135	0.0398 -		ISM02.4	PPB	7/30/2019 12:31	SMPL
K3877-04	MA52Q8	262	0.0706 -		ISM02.4	PPB	7/30/2019 12:33	SMPL
K3877-05	MA52S1	145	0.0422 -		ISM02.4	PPB	7/30/2019 12:36	SMPL
K3877-06	MA52M6D	162	0.0463 -		ISM02.4	PPB	7/30/2019 12:38	SMPL
K3877-07	MA52M6S	4894	1.1953 -		ISM02.4	PPB	7/30/2019 12:40	SMPL
PB121830BL	PBW830	132	0.039 -		ISM02.4	PPB	7/30/2019 12:43	SMPL
K3925-01	MA52J2	136	0.04 -		ISM02.4	PPB	7/30/2019 12:45	SMPL
K3925-02	MA52J2D	149	0.0431 -		ISM02.4	PPB	7/30/2019 12:47	SMPL
K3925-03	MA52J2S	4897	1.196 -		ISM02.4	PPB	7/30/2019 12:49	SMPL
K3925-05	MA52Q6	155	0.0446 -		ISM02.4	PPB	7/30/2019 12:52	SMPL
CCV010	CCV010	20567	5.0008 -		ISM02.4	PPB	7/30/2019 12:54	SMPL
CCB010	CCB010	101	0.0315 -		ISM02.4	PPB	7/30/2019 12:56	SMPL