

LB104309


7471B

INSTRUMENT ID: CU1

Linear

 μ Abs.:

57632



A= 0.0000e+000

B= 1.7291e-004

C= -7.0325e-002

Rho= 0.9998346

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	244	0.000	244				
0.2	0.200	0.167	-0.033	1371	0.0 %	1371				
2.5	2.500	2.500	-0.000	14863	0.0 %	14863				
5.0	5.000	5.088	0.088	29832	0.0 %	29832				
7.5	7.500	7.579	0.079	44242	0.0 %	44242				
10.0	10.000	9.895	-0.105	57632	0.0 %	57632				

LB104309 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 30 Jul 2019 13:15:58					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	244	-	0.00007471B	PPB		30 Jul 2019 13:19:43	S	Std
0.2 - 1	50-2	1371	-	0.20007471B	PPB		30 Jul 2019 13:22:01	S	Std
2.5 - 1	52-5	14863	-	2.50007471B	PPB		30 Jul 2019 13:24:20	S	Std
5.0 - 1	55-0	29832	-	5.00007471B	PPB		30 Jul 2019 13:26:38	S	Std
7.5 - 1	57-5	44242	-	7.50007471B	PPB		30 Jul 2019 13:28:55	S	Std
10.0 - 1	510-0	57632	-	10.00007471B	PPB		30 Jul 2019 13:31:12	S	Std
ICV52 - 1	ICV52	23165	3.9350	-7471B	PPB		30 Jul 2019 13:35:46	U	SMPL
ICB52 - 1	ICB52	171	-0.0408	-7471B	PPB		30 Jul 2019 13:38:02	U	SMPL
CCV11 - 1	CCV11	29685	5.0624	-7471B	PPB		30 Jul 2019 13:40:18	U	SMPL
CCB11 - 1	CCB11	165	-0.0418	-7471B	PPB		30 Jul 2019 13:42:33	U	SMPL
CRA - 1	CRA	1386	0.1693	-7471B	PPB		30 Jul 2019 13:44:50	U	SMPL
HighStd - 1	HighStd	57161	9.8132	-7471B	PPB		30 Jul 2019 13:47:06	U	SMPL
ChkStd - 1	ChkStd	41027	7.0235	-7471B	PPB		30 Jul 2019 13:49:22	U	SMPL
PB121831BL - 1	PBS	237	-0.0293	-7471B	PPB		30 Jul 2019 13:54:49	U	SMPL
PB121831BS - 1	LCSS	21099	3.5778	-7471B	PPB		30 Jul 2019 13:57:05	U	SMPL
K3621-07 - 1	OK-01-072919	2318	0.3305	-7471B	PPB		30 Jul 2019 13:59:22	U	SMPL
K3621-07DUP - 1	OK-01-072919DUP	2277	0.3234	-7471B	PPB		30 Jul 2019 14:01:40	U	SMPL
K3621-07MS - 1	OK-01-072919MS	23593	4.0091	-7471B	PPB		30 Jul 2019 14:03:57	U	SMPL
K3621-07MSD - 1	OK-01-072919MSD	23616	4.0130	-7471B	PPB		30 Jul 2019 14:06:15	U	SMPL
K3622-13 - 1	OR-03-072919-A	1718	0.2267	-7471B	PPB		30 Jul 2019 14:08:31	U	SMPL
CCV12 - 1	CCV12	29594	5.0467	-7471B	PPB		30 Jul 2019 14:10:47	U	SMPL
CCB12 - 1	CCB12	163	-0.0421	-7471B	PPB		30 Jul 2019 14:13:03	U	SMPL
K3622-15 - 1	OR-03-072919-B	2112	0.2949	-7471B	PPB		30 Jul 2019 14:15:19	U	SMPL
K4021-01 - 1	TP-10	2790	0.4121	-7471B	PPB		30 Jul 2019 14:17:35	U	SMPL
K4021-05 - 1	TP-9	2560	0.3723	-7471B	PPB		30 Jul 2019 14:19:51	U	SMPL
K4021-09 - 1	TP-7	377466	HIGH	-7471B	PPB		30 Jul 2019 14:22:06	X	
K4021-13 - 1	TP-8-A	5575	0.8936	-7471B	PPB		30 Jul 2019 14:23:58	U	SMPL
PB121832BL - 1	PBS	195	-0.0366	-7471B	PPB		30 Jul 2019 14:26:17	U	SMPL
PB121832BS - 1	LCSS	21297	3.6121	-7471B	PPB		30 Jul 2019 14:28:35	U	SMPL
K4024-01 - 1	TP-8	27207	4.6339	-7471B	PPB		30 Jul 2019 14:30:52	U	SMPL
K4024-01DUP - 1	TP-8DUP	28031	4.7764	-7471B	PPB		30 Jul 2019 14:33:10	U	SMPL
K4024-01MS - 1	TP-8MS	48062	8.2399	-7471B	PPB		30 Jul 2019 14:35:27	U	SMPL
CCV13 - 1	CCV13	29714	5.0674	-7471B	PPB		30 Jul 2019 14:37:44	U	SMPL
CCB13 - 1	CCB13	162	-0.0423	-7471B	PPB		30 Jul 2019 14:40:01	U	SMPL
K4024-01MSD - 1	TP-8MSD	50951	8.7394	-7471B	PPB		30 Jul 2019 14:42:19	U	SMPL
K4024-05 - 1	TP-6-A	4400	0.6905	-7471B	PPB		30 Jul 2019 14:44:36	U	SMPL
K4034-01 - 1	FK-PS-01-054	2754	0.4059	-7471B	PPB		30 Jul 2019 14:46:53	U	SMPL
K4044-03 - 1	TP-6	2136	0.2990	-7471B	PPB		30 Jul 2019 14:49:09	U	SMPL
K4044-05 - 1	TP-5	27993	4.7698	-7471B	PPB		30 Jul 2019 14:51:27	U	SMPL
K4044-09 - 1	TP-1	10953	1.8235	-7471B	PPB		30 Jul 2019 14:53:43	U	SMPL
K4044-15 - 1	TP-4	71818	12.3475	-7471B	PPB		30 Jul 2019 14:56:01	U	SMPL
K4044-17 - 1	TP-3	6451	1.0451	-7471B	PPB		30 Jul 2019 14:58:18	U	SMPL
K4049-01 - 1	SP-1	553960	HIGH	-7471B	PPB		30 Jul 2019 15:00:38	X	
K4049-02 - 1	SP-2	549842	HIGH	-7471B	PPB		30 Jul 2019 15:02:29	X	
CCV14 - 1	CCV14	28623	4.8788	-7471B	PPB		30 Jul 2019 15:04:24	U	SMPL
CCB14 - 1	CCB14	109	-0.0515	-7471B	PPB		30 Jul 2019 15:06:43	U	SMPL
K4049-03 - 1	SP-3	1037211	HIGH	-7471B	PPB		30 Jul 2019 15:09:00	X	
K4049-04 - 1	SP-4	302380	52.2131	-7471B	PPB		30 Jul 2019 15:10:51	U	SMPL
K4049-05 - 1	SP-5	110376	19.0144	-7471B	PPB		30 Jul 2019 15:13:19	U	SMPL
K4049-06 - 1	SP-6	176130	30.3837	-7471B	PPB		30 Jul 2019 15:23:23	U	SMPL
K4049-07 - 1	SP-7	323685	HIGH	-7471B	PPB		30 Jul 2019 15:25:39	X	
K4049-08 - 1	SP-8	210297	36.2914	-7471B	PPB		30 Jul 2019 15:27:34	U	SMPL
K4049-09 - 1	SP-9	651703	HIGH	-7471B	PPB		30 Jul 2019 15:29:54	X	
K4049-10 - 1	SP-10	1063598	HIGH	-7471B	PPB		30 Jul 2019 15:31:48	X	
K4049-11 - 1	SP-11	402224	HIGH	-7471B	PPB		30 Jul 2019 15:33:46	X	
K4049-12 - 1	SP-12	449794	HIGH	-7471B	PPB		30 Jul 2019 15:35:58	X	
CCV15 - 1	CCV15	28545	4.8653	-7471B	PPB		30 Jul 2019 15:37:53	U	SMPL
CCB15 - 1	CCB15	84	-0.0558	-7471B	PPB		30 Jul 2019 15:40:12	U	SMPL
K3621-07LX5 - 1		548	0.0244	-7471B	PPB		30 Jul 2019 15:42:29	U	SMPL
K3621-07A - 1		23578	4.0065	-7471B	PPB		30 Jul 2019 15:44:45	U	SMPL
K4024-01LX5 - 1		5482	0.8775	-7471B	PPB		30 Jul 2019 15:47:01	U	SMPL
K4024-01A - 1		49655	8.5153	-7471B	PPB		30 Jul 2019 15:49:18	U	SMPL
K4021-09DLX10 - 1		57089	9.8007	-7471B	PPB		30 Jul 2019 15:51:35	U	SMPL
K4044-15DLX4 - 1		17848	3.0157	-7471B	PPB		30 Jul 2019 15:53:52	U	SMPL
K4049-01DLX10 - 1		96102	16.5463	-7471B	PPB		30 Jul 2019 15:56:12	U	SMPL
K4049-02DLX10 - 1		93888	16.1635	-7471B	PPB		30 Jul 2019 15:58:29	U	SMPL
K4049-03DLX10 - 1		326911	56.4547	-7471B	PPB		30 Jul 2019 16:00:49	U	SMPL
K4049-04DLX10 - 1		31683	5.4079	-7471B	PPB		30 Jul 2019 16:03:08	U	SMPL
CCV16 - 1	CCV16	28865	4.9206	-7471B	PPB		30 Jul 2019 16:05:29	U	SMPL
CCB16 - 1	CCB16	145	-0.0453	-7471B	PPB		30 Jul 2019 16:07:47	U	SMPL
K4049-05DLX5 - 1		22359	3.7957	-7471B	PPB		30 Jul 2019 16:10:04	U	SMPL
K4049-06DLX10 - 1		17831	3.0128	-7471B	PPB		30 Jul 2019 16:12:20	U	SMPL
K4049-07DLX10 - 1		45770	7.8436	-7471B	PPB		30 Jul 2019 16:14:38	U	SMPL
K4049-08DLX10 - 1		21224	3.5994	-7471B	PPB		30 Jul 2019 16:16:55	U	SMPL
K4049-09DLX10 - 1		119927	20.6658	-7471B	PPB		30 Jul 2019 16:19:13	U	SMPL
K4049-10DLX10 - 1		334561	57.7774	-7471B	PPB		30 Jul 2019 16:21:31	U	SMPL
K4049-11DLX10 - 1		61379	10.5425	-7471B	PPB		30 Jul 2019 16:23:50	U	SMPL
K4049-12DLX10 - 1		71048	12.2143	-7471B	PPB		30 Jul 2019 16:26:13	U	SMPL
K4049-01DL2X40 - 1		24613	4.1854	-7471B	PPB		30 Jul 2019 16:28:33	U	SMPL
K4049-02DL2X40 - 1		24178	4.1102	-7471B	PPB		30 Jul 2019 16:30:51	U	SMPL
CCV17 - 1	CCV17	29749	5.0735	-7471B	PPB		30 Jul 2019 16:33:08	U	SMPL

LB104309 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Jul 2019 13:15:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB17 - 1	CCB17	97	-0.0536		-7471B	PPB	30 Jul 2019 16:35:24	U	SMPL
K4049-03DL2X100 - 1		36184	6.1861		-7471B	PPB	30 Jul 2019 16:37:41	U	SMPL
K4049-09DL2X100 - 1		11843	1.9774		-7471B	PPB	30 Jul 2019 16:39:58	U	SMPL
K4049-10DL2X100 - 1		35370	6.0454		-7471B	PPB	30 Jul 2019 16:42:16	U	SMPL
K4049-11DL2X100 - 1		6026	0.9716		-7471B	PPB	30 Jul 2019 16:44:32	U	SMPL
K4049-11DL2X100 - 1	CCV18	6970	1.1348		-7471B	PPB	30 Jul 2019 16:46:50	U	SMPL
CCV18 - 1		29402	5.0135		-7471B	PPB	30 Jul 2019 16:49:08	U	SMPL
CCB18 - 1		3	-0.0698		-7471B	PPB	30 Jul 2019 16:51:25	U	SMPL