



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104312

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/30/2019	09:40
2	K4050-07		1	9	100.00	>212.0	>212.0	07/30/2019	10:05
3	K4050-07DUP		1	9	100.00	>212.0	>212.0	07/30/2019	10:35
4	K4050-11		1	8	100.00	>212.0	>212.0	07/30/2019	11:10
5	K4050-19		1	7	100.00	>212.0	>212.0	07/30/2019	11:40
6	K4050-27		1	8	100.00	>212.0	>212.0	07/30/2019	12:05
7	K4051-30		1	9	100.00	>212.0	>212.0	07/30/2019	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 104312

WorkList Name : FP73019 WorkList ID : 129183 Department : Wet-Chemistry Date : 07-30-2019 09:08:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/05/2019	Water	K4050-07 B	Flash Point	Cool 4 deg C	PSEG01	N11	COMP-1	07/29/2019	1010A
08/05/2019	Water	K4050-11 B	Flash Point	Cool 4 deg C	PSEG01	N11	COMP-2	07/29/2019	1010A
08/05/2019	Water	K4050-19 A	Flash Point	Cool 4 deg C	PSEG01	N11	TOTES	07/29/2019	1010A
08/05/2019	Water	K4050-27 B	Flash Point	Cool 4 deg C	PSEG01	N11	COMP-3	07/29/2019	1010A
08/05/2019	Water	K4051-30 B	Flash Point	Cool 4 deg C	PSEG01	N11	COMP-3	07/29/2019	1010A

Date/Time 07/30/19 09:15
 Received by: KS
 Relinquished by: CP

Date/Time 07/30/19 12:45
 Received by: KS
 Relinquished by: KS