



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104318

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/30/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.00	07/30/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.02	07/30/2019	17:02
4	ICV	1	Water	NA	NA	20.1	6.99	07/30/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	07/30/2019	17:06
6	K4066-18	1	Water	NA	NA	22.6	7.05	07/30/2019	17:08
7	K4066-18DUP	1	Water	NA	NA	22.7	7.06	07/30/2019	17:09
8	CCV2	1	Water	NA	NA	20.1	12.02	07/30/2019	17:10

104318

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4067 V_A WorkList ID : 129204 Department : Wet-Chemistry Date : 07-30-2019 14:33:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/29/2019	Water	K4066-18-1	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-13	07/29/2019	9040C

Date/Time 07-30-19 16:00
Received by: JH
Relinquished by: JH

Date/Time 07-30-19 18:00
Received by: JH
Relinquished by: JH