



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104320

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/30/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.00	07/30/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.02	07/30/2019	17:02
4	ICV	1	Water	NA	NA	20.1	6.99	07/30/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	07/30/2019	17:30
6	K3618-10	1	Solid	20.02	20	26.1	8.12	07/30/2019	17:31
7	K4066-02	1	Solid	20.03	20	24.4	8.99	07/30/2019	17:32
8	K4066-08	1	Solid	20.04	20	24.6	9.89	07/30/2019	17:35
9	K4066-12	1	Solid	20.05	20	24.4	10.03	07/30/2019	17:36
10	K4066-16	1	Solid	20.04	20	24.6	9.77	07/30/2019	17:40
11	K4074-11	1	Solid	20.02	20	24.5	8.75	07/30/2019	17:41
12	K4074-12	1	Solid	20.03	20	24.6	8.72	07/30/2019	17:43
13	K4077-02	1	Solid	20.02	20	24.4	8.99	07/30/2019	17:45
14	K4077-02DUP	1	Solid	20.03	20	24.5	9.00	07/30/2019	17:46
15	CCV2	1	Water	NA	NA	20.1	12.02	07/30/2019	17:48

2-B 104320

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4066 WorkList ID : 129202 Department : Wet-Chemistry Date : 07-30-2019 14:32:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/30/2019	Solid	K3618-10 -1}	Corrosivity	Cool 4 deg C	PSEG05		HD-02-073019	07/30/2019	9045D
07/29/2019	Solid	K4066-02 -1}	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-1-A	07/29/2019	9045D
07/29/2019	Solid	K4066-08 -1}	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-11	07/29/2019	9045D
07/29/2019	Solid	K4066-12 -1}	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-12	07/29/2019	9045D
07/29/2019	Solid	K4066-16 -1}	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-13	07/29/2019	9045D
07/29/2019	Solid	K4074-11 -1}	Corrosivity	Cool 4 deg C	POWE02	O13	SS-12_072919	07/29/2019	9045D
07/29/2019	Solid	K4074-12 -1}	Corrosivity	Cool 4 deg C	POWE02	O13	SS-11_072919	07/29/2019	9045D
07/30/2019	Solid	K4077-02 -1}	Corrosivity	Cool 4 deg C	SEVE02	O21	20190612-LOC-13	07/30/2019	9045D

Date/Time 07-30-19 16:00
 Received by: JA
 Relinquished by: L.C

Date/Time 07-30-19 18:00
 Received by: SP
 Relinquished by: JA