

LB104321


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39277



A= 0.0000e+000

B= 2.5395e-004

C= -3.1094e-003

Rho= 0.9997622

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	1	0.000	1				
0.2	0.200	0.204	0.004	815	0.0 %	815				
2.5	2.500	2.409	-0.091	9499	0.0 %	9499				
5.0	5.000	5.166	0.166	20354	0.0 %	20354				
7.5	7.500	7.452	-0.048	29358	0.0 %	29358				
10.0	10.000	9.971	-0.029	39277	0.0 %	39277				

LB104321 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 31 Jul 2019 09:54:51				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50-0	1	-	0.00007470A	PPB	31 Jul 2019 10:17:33	S	Std
0.2 - 1	50-2	815	-	0.20007470A	PPB	31 Jul 2019 10:19:48	S	Std
2.5 - 1	52-5	9499	-	2.50007470A	PPB	31 Jul 2019 10:21:56	S	Std
5.0 - 1	55-0	20354	-	5.00007470A	PPB	31 Jul 2019 10:24:05	S	Std
7.5 - 1	57-5	29358	-	7.50007470A	PPB	31 Jul 2019 10:26:13	S	Std
10.0 - 1	510-0	39277	-	10.00007470A	PPB	31 Jul 2019 10:28:20	S	Std
ICV53 - 1	ICV53	15923	4.0406	-7470A	PPB	31 Jul 2019 10:30:51	U	SMPL
ICB53 - 1	ICB53	-39	-0.0130	-7470A	PPB	31 Jul 2019 10:32:57	U	SMPL
CCV19 - 1	CCV19	19500	4.9490	-7470A	PPB	31 Jul 2019 10:35:05	U	SMPL
CCB19 - 1	CCB19	-21	-0.0084	-7470A	PPB	31 Jul 2019 10:37:11	U	SMPL
CRA - 1	CRA	787	0.1968	-7470A	PPB	31 Jul 2019 10:39:17	U	SMPL
HighStd - 1	HighStd	39087	9.9232	-7470A	PPB	31 Jul 2019 10:41:24	U	SMPL
ChkStd - 1	ChkStd	26981	6.8488	-7470A	PPB	31 Jul 2019 10:43:31	U	SMPL
PB121858BL - 1	PBW	3	-0.0023	-7470A	PPB	31 Jul 2019 10:45:39	U	SMPL
PB121858BS - 1	LCSW	16154	4.0993	-7470A	PPB	31 Jul 2019 10:47:48	U	SMPL
K3621-08 - 1	OK-01-072919	12	-0.0001	-7470A	PPB	31 Jul 2019 10:49:56	U	SMPL
K3621-08DUP - 1	OK-01-072919DUP	45	0.0083	-7470A	PPB	31 Jul 2019 10:52:04	U	SMPL
K3621-08MS - 1	OK-01-072919MS	16350	4.1490	-7470A	PPB	31 Jul 2019 10:54:12	U	SMPL
K3621-08MSD - 1	OK-01-072919MSD	16120	4.0906	-7470A	PPB	31 Jul 2019 10:56:21	U	SMPL
K3622-14 - 1	OR-03-072919-A	24	0.0030	-7470A	PPB	31 Jul 2019 10:58:27	U	SMPL
CCV20 - 1	CCV20	19715	5.0036	-7470A	PPB	31 Jul 2019 11:00:34	U	SMPL
CCB20 - 1	CCB20	5	-0.0018	-7470A	PPB	31 Jul 2019 11:02:41	U	SMPL
K3622-16 - 1	OR-03-072919-B	39	0.0068	-7470A	PPB	31 Jul 2019 11:04:48	U	SMPL
K4044-04 - 1	TP-6	48	0.0091	-7470A	PPB	31 Jul 2019 11:06:55	U	SMPL
K4044-06 - 1	TP-5	61	0.0124	-7470A	PPB	31 Jul 2019 11:09:02	U	SMPL
K4044-10 - 1	TP-1	51	0.0098	-7470A	PPB	31 Jul 2019 11:11:08	U	SMPL
K4044-16 - 1	TP-4	59	0.0119	-7470A	PPB	31 Jul 2019 11:13:17	U	SMPL
K4044-18 - 1	TP-3	62	0.0126	-7470A	PPB	31 Jul 2019 11:15:25	U	SMPL
PB121859BL - 1	PBW	56	0.0111	-7470A	PPB	31 Jul 2019 11:17:33	U	SMPL
PB121859BS - 1	LCSW	16173	4.1041	-7470A	PPB	31 Jul 2019 11:19:41	U	SMPL
K4045-01 - 1	GW7-20190726-QC	18	0.0015	-7470A	PPB	31 Jul 2019 11:21:50	U	SMPL
K4045-01DUP - 1	GW7-20190726-QCDUP	54	0.0106	-7470A	PPB	31 Jul 2019 11:23:57	U	SMPL
CCV21 - 1	CCV21	19374	4.9170	-7470A	PPB	31 Jul 2019 11:26:03	U	SMPL
CCB21 - 1	CCB21	16	0.0010	-7470A	PPB	31 Jul 2019 11:28:10	U	SMPL
K4045-01MS - 1	GW7-20190726-QCMS	16093	4.0838	-7470A	PPB	31 Jul 2019 11:30:16	U	SMPL
K4045-01MSD - 1	GW7-20190726-QCMSD	16461	4.1772	-7470A	PPB	31 Jul 2019 11:32:23	U	SMPL
K4045-02 - 1	GW7-20190726	22	0.0025	-7470A	PPB	31 Jul 2019 11:34:30	U	SMPL
K4045-03 - 1	GW8-20190726	62	0.0126	-7470A	PPB	31 Jul 2019 11:36:37	U	SMPL
PB121800BL - 1	PBW	52	0.0101	-7470A	PPB	31 Jul 2019 11:38:44	U	SMPL
PB121800BS - 1	LCSW	16181	4.1061	-7470A	PPB	31 Jul 2019 11:40:52	U	SMPL
K3591-09 - 1	MDL-MDL-WATER-03-QT3-2019	195	0.0464	-7470A	PPB	31 Jul 2019 11:49:09	U	SMPL
PB121858TB - 1	PB121802TB	64	0.0131	-7470A	PPB	31 Jul 2019 11:51:17	U	SMPL
PB121859TB - 1	PB121807TB	73	0.0154	-7470A	PPB	31 Jul 2019 11:53:25	U	SMPL
K3621-08LX5 - 1		63	0.0129	-7470A	PPB	31 Jul 2019 11:55:33	U	SMPL
CCV22 - 1	CCV22	20024	5.0821	-7470A	PPB	31 Jul 2019 11:57:42	U	SMPL
CCB22 - 1	CCB22	30	0.0045	-7470A	PPB	31 Jul 2019 11:59:57	U	SMPL
K3621-08A - 1		16405	4.1630	-7470A	PPB	31 Jul 2019 12:02:04	U	SMPL
K4045-01LX5 - 1		35	0.0058	-7470A	PPB	31 Jul 2019 12:04:11	U	SMPL
K4045-01A - 1		16292	4.1343	-7470A	PPB	31 Jul 2019 12:06:18	U	SMPL
CCV23 - 1	CCV23	19596	4.9734	-7470A	PPB	31 Jul 2019 12:08:25	U	SMPL
CCB23 - 1	CCB23	2	-0.0026	-7470A	PPB	31 Jul 2019 12:10:33	U	SMPL